



Running Contract Details	
Equipment Name	Sample mixer
Running Contract Valid Till	25-09-2026
Tender Ref No	KMSCL/EP/T523/1385/2024
Tendered Quantity	25
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Sample mixer <i>Model & Make : SEROROLL 6/ REMI ELEKTROTECHNIK</i>	23997.66 Incl.GST :18%	1679.84	25677.5
		23997.66	1679.84	25677.5
Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
	Sample mixer			
Labour	500.00	500.00	500.00	
Comprehensive	1,000.00	2,000.00	3,000.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Sample mixer

Technical Specification

Equipment: Sample Mixer

1. Should be able to mix any biological sample including blood, cell suspensions
2. Should have compact and ergonomic design.
3. Should be able to keeps the contents of closed tubes in gentle motion.
4. Should accommodate various sizes of tubes
5. Should be able to vary speed
6. Should have speed range from 1 – 100 rpm.
7. Should have indication of power and speed.
8. Should work with AC 110/ 220V, 50/60Hz
9. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid