

Running Contract Details	
Equipment Name	Flexible Naso-pharygo Laryngoscope
Running Contract Valid Till	18-01-2026
Tender Ref No	KMSCL/EP/T514(R)/46/2023
Tendered Quantity	10
Supplier Name	M/s Lakshmi Surgicals
GST No	32AACFL2603A1ZB
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
2nd Floor 12/307-C1 KSIE Building CIC Freight Station ICTT Road Anavathil Jn. Udyogamandal Eloor Ernakulam - 683501	Contact Person	Mr. Viji Vijayan
	Phone	0484-2988006
	Mobile No	99466 02090
	Email	lakshmisurgicals2005@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Flexible Video Endoscope <i>Model & Make : 11102CMK/ Karl Storz GmbH & Co Germany</i>	713713.93 Incl.GST :12%	52636.4	766350.33
2	Rolling Stand for Monitor <i>Model & Make : 8404YA</i>	131243.01 Incl.GST :18%	9187.01	140430.02
3	Holder for flexible Endoscopes <i>Model & Make : 29005IFH</i>	53576.26 Incl.GST :18%	3750.34	57326.6
4	Monitor for endoscope <i>Model & Make : 8404ZXX</i>	726495.62 Incl.GST :28%	46881.67	773377.29
		1625028.82	112455.42	1737484.24

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Flexible Video Endoscope							
Labour	29,095.00	31,131.00	33,310.00	35,643.00	38,137.00	40,807.00	43,663.00
Comprehensive	1,78,998.00	1,91,527.00	2,04,933.00	2,19,280.00	2,34,628.00	2,51,053.00	2,68,627.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Flexible Video Endoscope

Equipment: Nasopharyngolaryngoscope

Flexible video Nasopharyngolaryngoscope required with video illumination to visualize and document the operational area for both adult & pediatric patient on video monitor. It should consist of following features

I. Video processing & Monitor

1. Monitor Screen size between 8 to 12 inch for display with touch screen to control features with HDMI output for connecting to a big screen which can display picture simultaneously on both screens.
2. Monitor should have two ports to connect two Video Nasopharyngolaryngoscope at one time and picture can be swapped using touch screen.
3. Monitor should be chargeable, to be supplied with charger and should have facility to be used while charging.
4. Monitor resolution should be minimum 1920 X 1200 pixels in 16:9 format.
5. It should have integrated video processing & Integrated recording of Video & still images should be possible on data card or USB drive with JPEG and MPEG format which can be easily transferred to the computer/laptop. Documented videos & still images should be easily recalled on the monitor.
6. Monitor should have a facility to connect flexible videoscopes directly without any special coupler or accessory.
7. Monitor should have Picture-in-Picture & side-by-side mode to view images from 2 different flexible video scopes.
8. Monitor should be splash proof according to IP 54 and should be shock resistant.
9. Monitor should have lithium-Ion rechargeable batteries and run for at least 100 minutes when fully charged.
10. Soft bag from same manufacturer should be supplied to place the monitor and system can also be operated without taking monitor out from the bag.

II. Flexible Video Endoscope

1. Flexible scope should be of chip-on-tip CMOS technology
2. Direction of view should be 0°
3. Angle of view should be 100°
4. Working length should be 29-31cm
5. Distal end outer diameter should be 2.8mm to 3.0mm
6. Deflection should be up to 140° up and 140° down
7. Following accessories should be included:
 - a. Case
 - b. Pressure Compensation Cap
 - c. Leakage Tester

III. All units should be provided from same manufacturer.

IV. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) Certification.

Note:

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail