



Running Contract Details	
Equipment Name	Autolensometer
Running Contract Valid Till	03-06-2026
Tender Ref No	KMSCL/EP/T519(R)/644/2023
Tendered Quantity	10
Supplier Name	M/s LUMEN MEDIWORKS PVT LTD
GST No	32AAECL4197J1ZB
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
63/1066A1 AISWARYAMRUTHAM CHEMMATH ROAD GANDHI NAGAR-682017	Contact Person	ARAVIND SUKUMAR
	Phone	0484-3551448
	Mobile No	6238826591
	Email	info@lumenmediworks.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Autolensometer <i>Model & Make : TL 6800/NINGBO MINSING</i>	95999.68 Incl.GST :12%	7079.98	103079.66
		95999.68	7079.98	103079.66

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Autolensometer							
Labour	1,944.00	1,944.00	1,944.00	1,944.00	1,944.00	1,944.00	1,944.00
Comprehensive	6,800.00	6,800.00	7,800.00	8,900.00	9,900.00	10,600.00	11,900.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Autolensometer

1. Should measure sphere range from -25 to +25
2. Should measure cylinder range from 0 to $\pm 10D$
3. Should measure Axis range from 1 to 180 degree in steps of 1 degree .
4. Should measure Prism range from 0 to 10
5. Should measure addition range from 0 to 10D
6. Should have a wave length of 535 to 660nm
7. Should measure lenses like glass lens, hard/soft contact lenses
8. Should have lens size maximum 10mm
 - a. Should have spectacle lens size 20 to 108 mm
9. Should have minimum 3.5 inch LCD/ LED colour monitor with touch screen. Should have memory function too.
10. Should be supplied with thermal printer/ink type
11. Should operate from 200 to 240v/50Hz Power supply
12. Should have UV tester, PD measuring, Auto progressive Lens measuring, distortion check
13. Should have safety certificate from a competent authority CE/FDA(US)/STQC CB certificate/STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid.

Note:

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail