



Running Contract Details	
Equipment Name	Electric Dermatone (Battery Driven)
Running Contract Valid Till	19-03-2025
Tender Ref No	KMSCL/EP/T482/874C/2022
Tendered Quantity	5
Supplier Name	M/s Praxis Health Technologies
GST No	32AAPFP1540D1ZM
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Puthiyapalam Tali Kozhikode Chalappuram P.O	Contact Person	Anoop M
	Phone	4952700425
	Mobile No	9747275757
	Email	praxis.health2013@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Electric Dermatone (Battery Driven)</b> <i>Model &amp; Make : B80D/G.R.Bioure Surgical System Pvt Ltd</i>	669760 Incl.GST :12%	49394.8	719154.8
		<b>669760</b>	<b>49394.8</b>	<b>719154.8</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Electric Dermatone (Battery Driven)</b>							
Labour	12,000.00	12,000.00	12,000.00	12,000.00	12,000.00	12,000.00	12,000.00
Comprehensive	31,395.00	32,964.00	34,613.00	36,344.00	38,161.00	40,069.00	42,072.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Electric Dermatone (Battery Driven)**

1. Cordless battery operated & light weight design
2. Precise thickness of the graft from 0.0 to 1, 2mm in 0.1 mm increments
3. Should have width reducing clamps 65, 50 and 35 mm to allow cutting of smaller graft widths
4. Battery and motor of the instrument are not sterilized
5. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate /STQC S certificate or valid detailed electrical and functional safety test report from ERTL.
6. Should have Standard cutting width- 80mm
7. Weight motor cartridge-325g (11.5 oz)
8. Motor capacity/rpm-1500/7.030min-1
9. Battery chemistry Li ion with no memory effect
10. Charger power supply-primary 100-240v/50-60Hz