



Running Contract Details	
Equipment Name	Cryobath
Running Contract Valid Till	03-08-2025
Tender Ref No	KMSCL/EP/T489(R)/153/2022
Tendered Quantity	15
Supplier Name	M/s Remi Elektrotechnik Ltd
GST No	27AAACR2578P1ZR
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
REMI House 3rd Floor 11 Cama Industrial Estate Walbhat Road Goregaon (East>. MumbaiÃ,Â. 400063	Contact Person	Mr Somyakanthya Pradhan
	Phone	022-40589820, 022-40589888
	Mobile No	9892604320
	Email	bids@remilabworld.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Cryobath <i>Model & Make : CB-704 with VS 03/ Remi Elektrotechnik Limited</i>	142240 Incl.GST :12%	10490.2	152730.2
		142240	10490.2	152730.2

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Cryobath				
Labour	3,300.00	3,300.00	3,300.00	3,300.00
Comprehens ive	4,600.00	4,600.00	4,600.00	4,600.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Cryobath

Equipment: Cryo Bath

1. Operating temperature: 3.80C to 4.20C
2. Programmable temp. range: 30C to 400C
3. Capacity: 12 regular plasma filled bags
4. Tray: Removable type SS trays partitions for holding bags
5. Time taken for one process: 5-6 hours (for plasma bags stored at - 400C)
6. Input supply: 230+/-10% VAC, 50 Hz

Note: If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid