



Running Contract Details	
Equipment Name	Water Bath Large
Running Contract Valid Till	03-10-2024
Tender Ref No	KMSCL/EP/T400/1169/2021
Tendered Quantity	500
Supplier Name	M/s Labline Instruments
GST No	32AABFL5818N1Z6
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
P B No. 36 Eroor Road Trippunithura-682301	Contact Person	Bini
	Phone	04842776582
	Mobile No	9020304422
	Email	lablinecochin@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Water Bath Large</b> <i>Model &amp; Make : SWB-4/ Labline</i>	13570 Incl.GST :18%	949.9	14519.9
		<b>13570</b>	<b>949.9</b>	<b>14519.9</b>
Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year	
<b>Water Bath Large</b>				
Labour	1,000.00	1,000.00	1,000.00	
Comprehensive	2,000.00	2,500.00	3,000.00	

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 3.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Water Bath Large**

#### **Water Bath Large**

1. Should have a double-walled construction.
  1. The inner chamber and top lid should be made of stainless steel.
  2. The space between the two walls should be packed with thick glass wool.
  3. Should provide with a microprocessor-based digital temperature controller with digital display
  4. The working temperature should be from ambient+5 °C to 80°C having an accuracy of +/- 0.5°C.
  5. Should have an approximate inner chamber dimension of 455mm x 300mm x 175mm
  6. Should have a water drainage facility
  7. Should have safety certificate from a competent authority CE/FDA (US). Other equivalent certificates will not be accepted.  
Should have valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid