

Running Contract Details	
Equipment Name	Embalming table
Running Contract Valid Till	22-02-2026
Tender Ref No	KMSCL/EP/T517(R)/1661/2023
Tendered Quantity	5
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Embalming table <i>Model & Make : CPHE66 /M/S. CAPT STEEL EQUIPMENT & FURNITURE</i>	47082 Incl.GST :18%	3295.74	50377.74
		47082	3295.74	50377.74

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Embalming table				
Labour	1,000.00	1,000.00	1,000.00	1,000.00
Comprehensive	3,000.00	4,000.00	5,000.00	5,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Embalming table

1. Should have Strong nonmagnetic stainless steel material frame work
2. Top is made with stainless steel sheet, with a high boarder
3. Provision of drainage hole, connected to a waste pipe
4. Bottom of the table covered with SS tube rail
5. Mounted on 125 mm swivel castors, two with breaks
6. Overall size: 190 (L) x 60(W) x 90(H) cm approximately

Note:

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail