



Running Contract Details	
Equipment Name	Microdrills with burr-heads (set)
Running Contract Valid Till	24-11-2024
Tender Ref No	KMSCL/EP/T460/72A1/2022
Tendered Quantity	20
Supplier Name	M/s Trivandrum Surgicals and Lab Supplies
GST No	32ERJPS6054N1Z5
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	5 Years

Supplier`s Details		
Address	Contact Details	
T.C.2/1094(7) SREELAKSHMI PAZHAYA ROAD MEDICAL COLLEGE P.O TVM.695015	Contact Person	GEETHA . S
	Phone	0471 - 2555040
	Mobile No	09447340147
	Email	trivandrumsurgicals@gmail.com

Item-wise Price Details							
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	<b>Microdrills with burr-heads (set)</b> <i>Model &amp; Make : Marathon / Saeyang</i>	235000.01 Incl.GST :12%	17331.25	252331.26			
		<b>235000.01</b>	<b>17331.25</b>	<b>252331.26</b>			
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year	11 <sup>th</sup> Year	12 <sup>th</sup> Year
<b>Microdrills with burr-heads (set)</b>							
Labour	8,832.00	8,832.00	8,832.00	8,832.00	8,832.00	8,832.00	8,832.00
Comprehensive	8,500.00	8,500.00	8,500.00	8,500.00	8,500.00	8,500.00	8,500.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Microdrills with burr-heads (set)**

1. Micromotor system should have the following specifications:
  - a) Speed Variation: Maximum should not be less than 30,000 rpm
  - b) Rotations: Clockwise & Anticlockwise
  - c) Foot Control: Variable speed type
  - d) Display: Digital display of rpm
2. Should have surgical type hand pieces suitable for ENT - Straight type & angled type
3. Set of Burrs suitable for the drill: Tungston Carbide Cutting and Diamond Polishing Burrs. (size 2mm, 3mm, 4mm, 5mm, 6mm, 7mm cutting and polishing) should be supplied