



Running Contract Details	
Equipment Name	MICROSCOPE - VARIOUS TYPES
Running Contract Valid Till	05-12-2024
Tender Ref No	KMSCL/EP/T453/13H/2022
Tendered Quantity	Item wise mentioned in below
Supplier Name	M/s New India Surgicals and Chemicals
GST No	32AAGFN8445E1ZD
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
30/1212IV Prasanth Buildings Near Medical College Hospital Calicut-673008	Contact Person	Prasanth N
	Phone	0495-2354188
	Mobile No	9447054188
	Email	nisc2000@rediffmail.com

Item-wise Price Details					
#	Item Details	Tendered Qty	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Dissecting Microscope</b> <i>Model &amp; Make : OX44/i7 Opto</i>	250	14632 <small>Incl.GST :18%</small>	1024.24	15656.24
2	<b>Binocular Microscope for Students</b> <i>Model &amp; Make : iOx 700 Bino/i7 Opto</i>	250	16081.04 <small>Incl.GST :18%</small>	1125.67	17206.71
			<b>30713.04</b>	<b>2149.91</b>	<b>32862.95</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year	
<b>Binocular Microscope Type IV</b>				
Labour	200.00		200.00	250.00
Comprehensive	350.00		400.00	450.00
<b>Dissecting Microscope</b>				
Labour	150.00		150.00	200.00
Comprehensive	300.00		350.00	400.00

<b>Annual / Comprehensive Maintenance Charges (Exl.Tax)</b>			
ve			
<b>Binocular Microscope for Students</b>			
Labour	150.00	150.00	200.00
Comprehensive	300.00	350.00	400.00
ve			
<b>Clinical Microscope</b>			
Labour	2,500.00	3,800.00	5,000.00
Comprehensive	5,000.00	7,500.00	10,000.00
ve			
<b>Microscope - LED Unicolor</b>			
Labour	150.00	150.00	200.00
Comprehensive	300.00	350.00	400.00
ve			

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Dissecting Microscope**

##### **Dissecting Microscope**

1. Stand : Dual light stable stand with Ball bearing focusing controls
2. Viewing Head: Binocular tube, 45degree inclined, 360 degree rotatable, IPD 55-75 mm, with Dioptic adjustment on left ocular
3. Eyepiece: Eyepiece Wide Field eyepiece,WF 10x/20mm paired , Revolving turret with objectives 2x - 4x
4. Magnification: 20x/40x
5. Illumination: Built-in Illumination system, Both LED incident light and sub-stage Lights for better vision.
6. Power:Power Input 220V AC.
7. Standard AccessoriesPacked in Molded Styrofoam box with: Operation Manual,Cleaning Cloth, Dust Cover
8. The product should be CE or FDA or BIS

#### **Equipment :Binocular Microscope for Students**

Ãçâ, -â€¹

## **Binocular Microscope for Students**

1. Inter changeable, inclined Binocular body, 360 degree rotatable head.
2. Eyepieces - Highest quality 10 X wide field anti fungus field eyepiece.
3. Objectives- Par focal, antifungus coated 4x, 10x, 40x and 100x (oil immersion) with semiplan achromatic correction for FOV of 18 mm.
4. Sub stage - Abbe condenser N.A 1.25 focusable, continuously variable iris diaphragm
5. Illuminator - Built-in LED/ Halogen light source with white light.
6. Finish - A durable textured acid resistant finish.
7. Should provide dust cover and immersion oil.
8. Should have Electrical safety certification
9. Should work with input 200 to 240Vac 50 Hz supply
10. The product should be CE or FDA or BIS Certified