



Running Contract Details	
Equipment Name	Automated Culture Identification System
Running Contract Valid Till	04-04-2025
Tender Ref No	KMSCL/EP/T479(R)/998A1/2022
Tendered Quantity	6
Supplier Name	M/s ABC Diagnostics Pvt Ltd
GST No	32AABCA8809P1ZE
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Njallanattu Buildings - S.H. Mount P.O Kottayam -16 Kerala	Contact Person	C.K.Suresh
	Phone	0481 -2565313
	Mobile No	9400865313 , 9447052281
	Email	abcdiagnostics.india@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Automated Culture Identification System <i>Model & Make : VITEK 2 Compact 60/ BioMerieux India Pvt Ltd</i>	1531999.9 Incl.GST :18%	107239.99	1639239.89
2	Rate of one card panel for ID GMP	262 Incl.GST :12%	19.32	281.32
3	Rate of one card panel for ID GPC	262 Incl.GST :12%	19.32	281.32
4	Rate of one card panel for SENSITIVITY GMP	262 Incl.GST :12%	19.32	281.32
5	Rate of one card panel for SENSITIVITY GPC	262 Incl.GST :12%	19.32	281.32
6	Rate of one card panel for Yeast	262 Incl.GST :12%	19.32	281.32
		1533309.9	107336.61	1640646.51

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Automated Culture Identification System							
Labour	55,000.00	55,000.00	55,000.00	55,000.00	55,000.00	55,000.00	55,000.00
Comprehensive	1,25,000.00	1,25,000.00	1,25,000.00	1,25,000.00	1,25,000.00	1,25,000.00	1,25,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Automated Culture Identification System

1. The system should be fully automated, walk away system for identification and antibiotic susceptibility testing for bacterial isolates, Identification of yeast and yeast like organisms
2. The system should be capable of simultaneous testing of minimum of 50 samples(25 identification and 25 antibiotic susceptibility testing)
3. Should be able to identify and perform AST for Gram positive bacteria, Gram negative bacteria.
4. The system should be able to detect antibiotic resistant organism like MRSA, VRE, HLAB, VRSA, B-lactamase and ESBL production
5. It should be an intelligent system and should give alert for any unusual antimicrobial resistance
6. The system should have barcode scanning system for easy management of samples
7. The system should be maintenance free without any need for regular calibrations, controls or standards run by the user
8. The system should use leak proof and non invasive system to avoid contamination of equipment and the environment
9. The identification system should complete in itself without the need of an additional test done manually
10. The system should have panels for identification alone or antibiotic susceptibility alone
11. The reagents/strips should have high stability and long shelf life
12. The system should have facilities for data management and storage and quality control
13. The system should be supplied in a complete system with all accessories (including ID Broth, AST Broth, AST indicator, Nephelometric Calibrator, Micropipette – 2 Nos, 1ml tip – 1 box, Vortex machine) , hardware's like computer, printer etc and the required software
14. The system should have expert software for analyzing the raw data and provide detailed interpretive results
15. Any software or database updates should be done free of cost by the firm during the life of the equipment as and when it is released by the manufacturer