



Running Contract Details	
Equipment Name	Disposable video Ureteroscope
Running Contract Valid Till	02-03-2025
Tender Ref No	KMSCL/EP/T479/270B/2022
Tendered Quantity	20
Supplier Name	M/s Medimart Medical Systems Pvt Ltd
GST No	33AAJCM1774D1ZN
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
9 Pillayar Koil Street Kaikankuppam Valasaravakkam Chennai 600087	Contact Person	Mohan Reddy
	Phone	4424400043
	Mobile No	9841043414
	Email	q.senthilkumar@gmail.com, ncmr4@yahoo.co.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Disposable video Ureteroscope (Separate order needs to be issued, if needed) <i>Model &amp; Make : HU30S / SZ HUGEMED TECH DEV CO LTD, CHINA</i>	61600 Incl.GST :12%	4543	66143
2	Disposable video Cystoscope (Separate order needs to be issued, if needed)	50400 Incl.GST :12%	3717	54117
3	Control Unit (Separate order needs to be issued, if needed)	224000 Incl.GST :12%	16520	240520
4	Grasping Forceps (Separate order needs to be issued, if needed)	8960 Incl.GST :12%	660.8	9620.8
5	Biopsy Forceps (Separate order needs to be issued, if needed)	8960 Incl.GST :12%	660.8	9620.8

Item-wise Price Details							
6	Access Sheath (Separate order needs to be issued, if needed)		5600 Incl.GST :12%	413	6013		
7	Pressure compensation cap, Leakage tester, Cleaning Brush, Carry Case (Separate order needs to be issued, if needed)		5600 Incl.GST :12%	413	6013		
			<b>365120</b>	<b>26927.6</b>	<b>392047.6</b>		
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Disposable video Ureteroscope (Separate order needs to be issued, if needed)</b>							
Labour	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00
Comprehensive	20,000.00	20,000.00	20,000.00	20,000.00	20,000.00	20,000.00	20,000.00

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment : Disposable video Ureteroscope (Separate order needs to be issued, if needed)**

#### **I. Disposable Video Ureteroscope**

1. Should be a flexible fiberoptic Uretero – Renoscope with Continuous Irrigation.
2. Outer diameter of the shaft should be between 6 Fr. - 10 Fr.
3. Should have an instrument channel between 3 Fr. to 4 Fr.
4. Should have 660mm ±20mm working length.
5. Should have a ceramic liner in the distal end of the working channel to protect it from thermal or electro Cautery damage.
6. Should have field of view between 80- 120 degrees, direction of view 0 degree.

7. Should have minimum angle of deflection 270 or more

8. Should have active deflection mechanism.

9. Should be supplied with the following accessories of compatible sizes.

a. Grasping Forceps

b. Biopsy Forceps

c. Access sheath

10. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid

## **II. Flexible Disposable Video Cystoscope**

1. Tip size of the flexible Cystoscope should be between 9Fr. to 18F.

2. Cystoscope Lateral deflection should be 180degree minimum on both sides.

3. Scopes life/duration should not be restricted with timer.

4. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid

## **III. Control Unit :**

1. Storage / Data Transfer: USB

2. Image Functions: White Balance, Brightness control, image capture, Video Capture

3. Video Connections : SDI, DVI & HDMI

4. Power : 100V AC to 240V AC

5. Frequency : 50Hz

6. Lamp : 60 W LED