



Running Contract Details	
Equipment Name	Procedure Table(Dressing)
Running Contract Valid Till	27-02-2026
Tender Ref No	KMSCL/EP/T516/22B/2023
Tendered Quantity	500
Supplier Name	M/s IMEDFURNS PRIVATE LIMITED
GST No	33AAGCI1214D1ZB
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
61/1A SERVICE ROAD PERIYAPANICHERRI CHENNAI -600128	Contact Person	MR.PRADEEP KUMAR
	Phone	
	Mobile No	9500120662
	Email	tenders@imedfurns.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	CLEANING AND DRESSING TABLE(PROCEDURE TABLE) <i>Model &amp; Make : IMED5705K / IMEDFURNS</i>	22656 Incl.GST :18%	1585.92	24241.92
		<b>22656</b>	<b>1585.92</b>	<b>24241.92</b>

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :CLEANING AND DRESSING TABLE(PROCEDURE TABLE)**

##### **Equipment: Cleaning And Dressing Table ( Procedure Table)**

1. Completely made of SS304 grade.
2. Should have a stainless steel top of grade SS 304 gauge 16.
3. Should be 7 feet long and 2 feet wide with four legs
4. Legs should be SS 304 with 2 inches square tube 18 gauge thickness or better.
5. Should have proper reinforcement for the top
6. Should be suitable for all type of treatment procedures in OP including fracture treatment in ortho OP.
7. Should be supplied with suitable foot step made of SS304.
8. Material test certificate for the proof of SS 304 shall be furnished along with the invoice and the manufacturer shall also submit undertaking that the tested material is used for the construction of the equipment supplied.

##### **Note:**

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail