



Running Contract Details	
Equipment Name	Oto Acoustic Emission Screeners Type B
Running Contract Valid Till	15-09-2023
Tender Ref No	KMSCL/EP/T404(R)/513B/2021
Tendered Quantity	100
Supplier Name	M/s Auditivo Hearing Services Private Limited
GST No	07AALCA7725L1Z8
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
309 Oriental House 19 Community Centre Yusuf Sarai New Delhi	Contact Person	Ajay Kuamr Mathur
	Phone	1149874415
	Mobile No	9810112042
	Email	auditivohearing@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Oto Acoustic Emission Screeners Type B</b> <i>Model &amp; Make : Path Sentiero/ path Med</i>	274050 Incl.GST :5%	21558.6	295608.6
		<b>274050</b>	<b>21558.6</b>	<b>295608.6</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Oto Acoustic Emission Screeners Type B</b>							
Labour	5,000.00	5,500.00	6,000.00	7,000.00	7,500.00	8,000.00	8,500.00
Comprehensive	9,000.00	9,000.00	9,000.00	10,000.00	10,000.00	10,500.00	10,500.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 3.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Oto Acoustic Emission Screeners Type B**

#### **Oto Acoustic emission screeners Type B**

1. System should be a handheld, lightweight, portable unit with color touch screen with Keyboard.
2. Should have AC line voltage of 220 V and 50 Hz along with the rechargeable inbuilt battery operated power supply
3. Should have long battery life with a minimum of 10 hrs of battery backup which can be chargeable through direct power.
4. Should have the facility to perform screening and diagnostic testing with TEOAE and DPOAE.
5. Should have DPOAE (1.5-8 KHz) or TEOAE (1-4 KHz) measurements on both ears
6. Should have test battery of both screening and /DPOAE-testing.
7. Should have the facility to edit protocol used- Customize measurement protocols and display parameters.
8. Should have a provision of Automatic pass/ refer indicators with each frequency SNR value in screening mode.
9. Should be fast, noise-resistant (High noise immunity) OAE recording with no electrical interference, preferably online noise indicators [desirable].
10. Should have a database with storing capacity of 150 test results.
11. Should have PC link software and data software.
12. Should have LED lights for stimulus, noise or low battery.
13. Should be supplied with carrying case
14. Should have probe mount kit (capable clip); probe Cleaning kit; Calibration/test cavity
15. Should have automatic shut down facility if not in use and LED indication/display for low battery.
16. Should have the facility to upgrade to ABR, ASR and higher version and to incorporate additional test if any for the same module (Hardware & Software).
17. Should have a safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid