



Running Contract Details	
Equipment Name	Automatic External Defibrillator (AED)
Running Contract Valid Till	22-06-2024
Tender Ref No	KMSCL/EP/T422(R)/648/2021
Tendered Quantity	100
Supplier Name	M/s Biocare Diagnostics
GST No	32AAOFB5438P1ZZ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
787a-1 First Floor Kamala Building Chalappuram (Po) Calicut 673002	Contact Person	Linesh.T
	Phone	
	Mobile No	9847009640
	Email	mail@biocarecalicut.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Automatic External Defibrillator (AED) as per the specification except the following <i>Model & Make : AED 3100/NIHON KOHDEN</i>	96768 Incl.GST :12%	7136.64	103904.64
2	AED Pad	3143.84 Incl.GST :12%	231.86	3375.7
		99911.84	7368.5	107280.34

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Automatic External Defibrillator (AED) as per the specification except the following							
Labour	4,500.00	4,500.00	4,500.00	4,500.00	4,500.00	4,500.00	4,500.00
Comprehensive	7,258.00	7,621.00	7,984.00	8,347.00	8,710.00	9,073.00	9,436.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 3.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Automatic External Defibrillator (AED) as per the specification except the following

ItemName: Automatic External Defibrillator (AED)

1. Should have maximum energy output of atleast 150J (biphasic low energy waveform) with charging time less than 10seconds.
2. Should be capable to use on adult and paediatric patients.
3. Should have CPR guidance facility.
4. Should work on pregelled pad electrodes same for both Adult and child patients.
5. Should have automatic/configurable self test facility.
6. Should be disposable or nonrechargeable batteries capable of delivering minimum 200 full energy shocks or at least 1.5hrs continuous monitoring or defibrillation.
7. Should have facility to record event data.
8. Should have audio visual prompts and alarms at various stages of operation.
9. Should be light weight, portable and should be supplied with a carrying case.
10. The rate of 50nos of disposable pad electrodes shall be quoted separately and this rate will be taken for evaluation.
11. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or validdetailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with thetechnical bid.
12. Should have 1 m drop test passed, along with at least IP 55 rating.