



Running Contract Details	
Equipment Name	Auotmated Rotary Microtome
Running Contract Valid Till	24-11-2024
Tender Ref No	KMSCL/EP/T454/626D/2022
Tendered Quantity	5
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details							
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	<b>Auotmated Rotary Microtome</b> <i>Model &amp; Make : YD-355AT/ Jinhua Yidi Medical</i>	736320 Incl.GST :18%	51542.4	787862.4			
		<b>736320</b>	<b>51542.4</b>	<b>787862.4</b>			
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Auotmated Rotary Microtome</b>							
Labour	10,000.00	10,000.00	10,000.00	12,000.00	12,000.00	12,000.00	15,000.00
Comprehensi ve	15,000.00	15,000.00	20,000.00	20,000.00	25,000.00	25,000.00	30,000.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Automated Rotary Microtome**

1. Ideally located coarse feed wheel for fatigue – free operation.
2. Extremely smooth-running hand wheel.
3. Safety locking mechanism on the hand wheel.
4. Section thickness setting range-0.5-60µm
5. Trimming thickness ranges – 10,50 µm
6. Total horizontal specimen feed – 25mm
7. Vertical stroke length – 59mm
8. Specimen retraction.
9. Specimen orientation – horizontal 8 degree & vertical 8 degree
10. Clause deleted
11. Single – handed operation specimen clamps
12. Facility to use cassette clamps should be available.
13. Precise knife holder lateral adjustments ensures that the entire width of the blade can be utilized.
14. Universal knife holder base for different knife holders(Disposable, Standard & Heavy duty).
15. Disposable blade holder with integrated safety finger protection guard.
16. Accessories to be supplied along with the machine:
  1. Cassette Clamp -1 No.
  2. Disposable Microtome Blades – 2 packets (50nos/pack)
  3. Dust Cover, Oil, Manual