

(A Government of Kerala Undertaking) Thycaud P.O, Thiruvananthapuram - 14, Kerala. Tel: 0471 - 2945600, 2337353, Fax: 0471 - 2945647 Email :ep.kmscl@kerala.gov.in CIN: U24233KL200TSGC021616, PAN : AADCK4029M, GSTIN : 32AADCK4029M1ZK

Running Contract Details			
Equipment Name	Nebulizer-Piston Compression		
Running Contract Valid Till	18-02-2024		
Tender Ref No	KMSCL/EP/T414/82/2021		
Tendered Quantity	500		
Supplier Name	M/s Kerala Surgical Equipment Co		
GST No	32AACFK7295D1ZG		
Installation & Delivery Period	6 Week(s)		
Up-time / PM vist	95% & 0 Visits per year		
Warranty period	1 Years		

Supplier`s Details					
Address	Contact Details				
No.41/1737 A	Contact Person	Mr. Abbas			
MES Centre Near Town Hall Banerji Road	Phone				
Cochin -18	Mobile No	9539012460			
	Email	info@keralasurgicals.com			

Item-wise Price Details						
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total		
1	Nebulizer-Piston Compression Model & Make : Lifeline	2100 Incl.GST :12%	154.88	2254.88		
		2100	154.88	2254.88		

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.

3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.

4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Nebulizer-Piston Compression

- 1. Should be Piston Compression Type
- 2. Should be lightweight, portable and compact.
- 3. Should have a dust filter.
- 4. Should be able to deliver a flow rate ? 7 lpm
- 5. Should have air pressure ? 16 psi.
- 6. Should have a check valve to protect the device against contamination due to backward inhalation
- 7. Should be compatible for continuous use
- 8. Should works on 200-240Vac/50Hz.
- 9. Should be supplied with nebulization accessory kit with mask for adult and paediatric -2 nos. each
- 10. Nebulization mask for adult and pediatric 10 nos. each
- 11. Should be heavy duty type Nebulizer suitable for usage in Govt. Hospitals