



Running Contract Details	
Equipment Name	HOSPITAL FURNITURE
Running Contract Valid Till	21-03-2024
Tender Ref No	KMSCL/EP/T414/328/2021
Tendered Quantity	200
Supplier Name	M/s Carevel Medical Systems Pvt Ltd
GST No	09AABCC6250C1Z1
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
S-35-36 Site- IV Suhibabad Industrial Area Ghaziabad - 201 010	Contact Person	Mr. Gyanendra Rajvanshi
	Phone	0120-4292637
	Mobile No	9599679916
	Email	rkurup@carevelglobal.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Dressing Trolley <i>Model & Make : C-4101</i>	17091.12 Incl.GST :18%	1196.38	18287.5
2	Patient Trolley	24367 Incl.GST :18%	1705.69	26072.69
		41458.12	2902.07	44360.19

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.

4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Dressing Trolley

1. Overall approximate dimension: 1000 mm L X 500 mm W X 900 mm H \pm 50 mm tolerance accepted.
2. Approximate shelf dimension: 750 mm L x 500 mm W.
3. S.S. tubular frame mounted on four 125 mm diameter castors with synthetic body, two with brake & two without brake.
4. Two S.S. shelves with protective railings on all four sides with 10mm SS rods.
5. The sheets used shall be of 1.2 mm thick.
6. With S.S. bowl and S.S. bucket
7. Note : Material test certificate for the proof of SS 304 shall be furnished along with the invoice and the manufacturer shall also submit undertaking that the tested material is used for the construction of the equipment supplied.

Equipment :Patient Trolley

1. Size : 2000mm (L) x 660mm (W) x 830mm (H)
2. Main frame : 32mm OD x 1.6 mm Thick Stainless steel grade SS 304
3. Supporting frame : 25mm OD x 1.6 mm Thick Stainless steel grade SS 304
4. 4 castor wheels with 200 mm dia.
5. S.S. tubular frame mounted on four 125 mm diameter castors with synthetic body, two with brake & two without brake
6. Stretcher top frame 25mm OD x 1.6 mm Thick Stainless steel grade SS 304
7. Sheet should be made of Stainless steel grade SS 304 of 1.2mm thickness
8. Provision for IV pole.
9. Max. Load of 200 Kg
10. Provision for carrying oxygen cylinder in Stainless steel grade SS 304
11. Finish : Stainless steel grade SS 304