



Running Contract Details	
Equipment Name	OT table Orthopedics - manual hydraulic
Running Contract Valid Till	20-10-2024
Tender Ref No	KMSCL/EP/T442/93/2022
Tendered Quantity	20
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>OT table Orthopedics - manual hydraulic</b> <i>Model &amp; Make : PSI VIRTUE ORTHO/ Palakkad Surgicals</i>	470820 Incl.GST :18%	32957.4	503777.4
		<b>470820</b>	<b>32957.4</b>	<b>503777.4</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>OT table Orthopedics - manual hydraulic</b>							
Labour	12,000.00	12,000.00	12,000.00	12,000.00	12,000.00	12,000.00	12,000.00
Comprehensive	31,900.00	31,900.00	31,900.00	31,900.00	31,900.00	31,900.00	31,900.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :OT table Orthopedics - manual hydraulic**

1. The table should have minimum 4 sections
2. The table should have a detachable split leg and detachable head section.
3. The table top should be completely radiolucent.
4. The table top should have a minimum height of 765mm.
5. The table top should have a minimum vertical stroke of 300mm
6. Should have at least 25° trendelenburg or reverse trendelenburg
7. Should have at least 20° lateral tilt movement
8. Should have at least 80° back up movement
9. The head section should have up and down movement.
10. The table should have a heavy and sturdy base and compact to provide adequate foot room for the operating team.
11. The table should be mounted on heavy duty casters which offer enhanced weight bearing capacity and free mobility.
12. The table should have a single lever foot operated brake pedal
13. Should have a minimum patient weight bearing capacity of 175 Kgs.
14. Base should be made of cast Iron and all other parts and accessories should be completely made up of stainless steel.
15. The size of the main central column should be smaller to offer higher c-arm access.
16. The table should be compatible for spine surgeries with C-arm or separate accessory should be provided for doing spine surgery with C-arm.
17. The pelvic support should be movable to the edges and completely radiolucent.
18. The perennial post should be completely radiolucent
19. The table should be supplied with the following accessories.
  - a) Mattress for the complete table top in sections - 1 set
  - b) A pair of arm boards with pad and fixing clamp - 1

- c) A pair of padded shoulder support with clamps - 1
- d) A pair of padded lateral support with clamps - 1
- e) A pair of leg crutches with clamps - 1
- f) Anesthetic screen frame with clamp - 1
- g) Adjustable traction device with minimum leg stroke of 175mm
- h) Should have accessories for tibia interlocking - upward and downward
- i) Telescopic Arm surgery board with floor support