

Running Contract Details	
Equipment Name	Muscle Stimulator Therapeutic Model
Running Contract Valid Till	05-01-2025
Tender Ref No	KMSCL/EP/T463/1315/2022
Tendered Quantity	100
Supplier Name	M/s Biocare Diagnostics
GST No	32AAOFB5438P1ZZ
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
787a-1 First Floor Kamala Building Chalappuram (Po) Calicut 673002	Contact Person	Linesh.T
	Phone	
	Mobile No	9847009640
	Email	mail@biocarecalicut.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Muscle Stimulator Therapeutic Model <i>Model & Make : DSL 007/ GALTRON ELECTRO MEDICAL PVT. LTD</i>	9296 Incl.GST :12%	685.58	9981.58
		9296	685.58	9981.58

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Muscle Stimulator Therapeutic Model

1. It should be a microprocessor based, therapeutic, muscle stimulator unit for muscle stimulation
2. It should be light weight and compact in design.
3. It should have visual/audio indicator for different mode of treatment.
4. It should have a soft touch key to select treatment modes.
5. Machine should produce output only after intensity is set to Zero with audio alarm
6. It should enable user to select accurate pulses with pulse width for treatment in fraction of seconds.
7. It should have digital display of output current to accurately set output current
8. It should have surge protection safety feature circuit to protect the machine and patient if high surge is sensed in the input current .
9. It should have microprocessor controlled digital display of treatment time. The output should switch off automatically after treatment time is over.
10. It should have Plain Galvanic, Pulsed Galvanic, Plain Faradic and Surged Faradic for muscle stimulation.
11. Pulseduration.01,.03,.1,.3,1,3,10,30,100,300 mSec.
12. Faradic frequency- 50 Hz/.7 mSec
13. Surge duration 1-6 sec
14. Output current 0-40 mA
15. Unit should function with 200-240 AC/ 50- 60 Hz input power supply.
16. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL.