



Running Contract Details	
Equipment Name	Tympanometer
Running Contract Valid Till	12-08-2023
Tender Ref No	KMSCL/EP/T404/1297/2021
Tendered Quantity	50
Supplier Name	M/s Auditivo Hearing Services Private Limited
GST No	07AALCA7725L1Z8
Installation & Delivery Period	
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
309 Oriental House 19 Community Centre Yusuf Sarai New Delhi	Contact Person	Ajay Kuamr Mathur
	Phone	1149874415
	Mobile No	9810112042
	Email	auditivohearing@gmail.com

Item-wise Price Details							
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	Tympanometer <i>Model & Make : MI34H Touchtym/MAICO Diagnostics</i>	292950 Incl.GST :5%	23045.4	315995.4			
		292950	23045.4	315995.4			
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Tympanometer							
Labour	5,500.00	6,500.00	7,500.00	7,500.00	8,500.00	8,500.00	8,500.00
Comprehensive	9,000.00	9,000.00	9,000.00	10,000.00	10,500.00	11,000.00	12,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment : Tympanometer

Tympanometer

1. Probe tone: 226 and 1000Hz (For Pediatric testing also)
2. Probe assembly with contralateral test facility with supra-aural earphones: (TDH39/ TDH39A/ TDH49/ TDH49A/ TDH 50 with MX 41 AR ear cushions) or insert earphones (ER Tone 3A)
3. Test cavities (0.5, 2, 5 cc)
4. Probe tips - assorted
5. Shall have Printer
6. Tests required
 - a. Compensated tympanometry (ear canal volume and tympanometric peak pressure)
 - b. Ipsilateral and contralateral acoustic reflexes
 - c. Eustachian tube function tests - intact and perforated
7. Air pressure range: + 200da Pa to – 400 da Pa
8. Stimuli for acoustic reflexes:
 - a. Type: Pure tones
 - b. Frequencies: 500Hz, 1000Hz, 2000Hz and 4000Hz
 - c. Intensity: up to 120 dB HL
9. Power supply: UPS/Battery backup & mains.
10. Should operate in 200-240 V AC 50 Hz
11. Shall have Self-calibration
12. Should have a safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
13. Annual Calibration shall be performed in warranty and CMC period.
14. Ear Tips should be assorted and supply Big Size - 5 Nos and all other sizes - 10 Nos
15. Probe Tips should be supplied - 2 Nos
16. Thermal Paper Roll - 5 Nos
17. Software for connecting to the PC along with necessary cables shall be provided.