



Running Contract Details	
Equipment Name	Oto Acoustic Emission Screeners Type B
Running Contract Valid Till	04-08-2026
Tender Ref No	KMSCL/EP/T539/513B/2024
Tendered Quantity	75
Supplier Name	M/s Medilife Technologies
GST No	07AAJFM2355B1ZL
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
1st Floor Plot No. 2 Vikas Plaza LSC Kalkaji New Delhi-110 019	Contact Person	Mr. Sanjay Sharma
	Phone	01140588607
	Mobile No	9810524302
	Email	info@medilife.in

Item-wise Price Details							
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	Oto Acoustic Emission Screeners Type B <i>Model & Make : ascreen/Neurosoft LIC</i>	246750 Incl.GST :5%	19411	266161			
		246750	19411	266161			
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Oto Acoustic Emission Screeners Type B							
Labour	4,700.00	4,700.00	4,700.00	4,700.00	4,700.00	4,700.00	4,700.00
Comprehen sive	9,500.00	9,500.00	9,500.00	9,500.00	9,500.00	9,500.00	9,500.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Oto Acoustic Emission Screeners Type B

Item Name: Oto Acoustic emission screeners Type B

1. System should be handheld, lightweight, portable unit with color touch screen with Key board.
2. Should have AC line voltage of 220 V and 50 Hz along with rechargeable inbuilt battery operated power supply
3. Should have long battery life with minimum of 10 hrs of battery backup which can be chargeable through direct power.
4. Should have facility to perform screening and diagnostic testing with TEOAE and DPOAE.
5. Should have DPOAE (1.5-8 KHz) or TEOAE (1-4 KHz) measurements on both ears
6. Should have test battery of both screening and /DPOAE-testing.
7. Should have the facility to edit protocol used- Customize measurement protocols and display parameters.
8. Should have provision of Automatic pass/ refer indicators with each frequency SNR values in screening mode.
9. Should be fast, noise resistant (High noise immunity) OAE recording with no electrical interference, preferably online noise indicators [desirable].
10. Should have data base with storing capacity of 150 test results.
11. Should have PC link software and data software.
12. Should have LED lights for stimulus, noise or low battery.
13. Should supplied with carrying case
14. Should have probe mount kit (capable clip); probe Cleaning kit; Calibration/test cavity
15. Should have automatic shut down facility if not in use and LED indication/display for low battery.
16. Should have facility to upgrade to ABR, ASRand higher version and to incorporate additional test if any for the same module (Hardware & Software).
17. Should have safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid