



Running Contract Details	
Equipment Name	Vaccum extractor
Running Contract Valid Till	21-08-2024
Tender Ref No	KMSCL/EP/T416/327A/2021
Tendered Quantity	500
Supplier Name	M/s C Cube Advanced Technologies
GST No	29ANCP4091F1ZS
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
406 3rd floor Malnad Residency 4th Main Hanuma Hills layout Banashankari 3rd stage Bangalore 560061	Contact Person	Vinayaka KS
	Phone	8043021028
	Mobile No	9611111273
	Email	ccubeadvtech@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Vaccum extractor <i>Model & Make : VAC ASSIST/C CUBE ADVANCED TECHNOLOGIES</i>	89600 Incl.GST :12%	6608	96208
		89600	6608	96208

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Vaccum extractor							
Labour	3,000.00	3,000.00	3,000.00	3,000.00	3,000.00	3,000.00	3,500.00
Comprehensive	6,000.00	6,000.00	6,200.00	6,200.00	6,200.00	6,400.00	6,400.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Vaccum extractor

1. Should have Vacuum generator, antibacterial plastic cups, silastic cup and metal cup of different sizes with necessary electrical connection cable
2. Should have cordless Foot Switch operation
3. Should generate a steady vacuum of at least 600mmHg
4. Should work on 200-240Vac 50Hz power supply
5. Should have digital display
6. Should have traction force indicating display in kilograms