



Running Contract Details	
Equipment Name	Multiparameter Monitor - Neonatal
Running Contract Valid Till	11-04-2026
Tender Ref No	KMSCL/EP/T518/75B/2024
Tendered Quantity	150
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details							
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	Multiparameter Monitor - Neonatal <i>Model & Make : XM750/ Shenzhen Hwatime Biological Medical Electronics Co., Ltd</i>	55372.8 Incl.GST :12%	4083.74	59456.54			
2	Neonatal ECG patient cable (Rate for future purchase)	3628.8 Incl.GST :12%	267.62	3896.42			
3	Neonatal SpO2 probe with cable (Rate for future purchase)	2788.8 Incl.GST :12%	205.67	2994.47			
4	Monitor stand including instalation charges (Rate for future purchase)	3012.8 Incl.GST :12%	222.19	3234.99			
		64803.2	4779.24	69582.44			
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Multiparameter Monitor - Neonatal							
Labour	900.00	900.00	900.00	900.00	900.00	900.00	900.00
Comprehensive	3,000.00	3,500.00	4,000.00	4,500.00	5,000.00	5,500.00	6,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Multiparameter Monitor - Neonatal

1. The system should be a dedicated neonatal multiparamater monitor
2. Should be able to monitor 6 channel ECG from 5 lead ECG, SPO2, NIBP, Respiration Rate and temperature.
3. Should be portable with carrying handle.
4. Should have touch screen TFT display with at least 12 inches or higher with at least 6 waveforms and numeric display simultaneously.
5. Should have Lithium ion battery with 3 hours battery back up.
6. Should have keys for quick access to main functions.
7. Should provide prominent prioritized audio, visual alarms for high, low heart rate, SpO2, RR, low battery and lethal arrhythmia recognition with detailed list (more than 15 parameters) in the monitor to set priority and ST Analysis.
8. NIBP can be taken on manual/auto/stat modes.
9. The following accessories should be supplied along with machine free of cost:
 - a. Reusable neonatal 5 lead ECG cable set – 2 nos.
 - b. Reusable neonatal SPO2 finger probes – 2 nos.
 - c. NIBP cuff of 5 different sizes
10. Should work on 200-240V AC/50Hz with inbuilt rechargeable battery.
11. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
12. Should incorporate motion tolerance capability and low perfusion index. Copy of validation proof of these two features from the manufacturer should be submitted in the technical bid and during demonstration
13. Monitor should have networking facility with bidirectional & bed to bed communication.
14. Monitor stand should be supplied along with the machine as per the following specification

- a. Monitor stand -extruded Aluminum, powder coated.
- b. Load bearing capacity 20 kgs approximately.
- c. Should supply, install with necessary anchor fastners at the site.
- d. Equipment performance should not be affected by electromagnetic interferences radiated or conducted through power lines from another device.

Note: :

1. The cost of probes and monitor stand can be quoted separately for future purchases and will be considered for L1 calculation. The price shall be fixed for 3 years from the date of price bid opening.
2. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
3. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail