



Running Contract Details	
Equipment Name	Litho Clast Pneumatic Ultrasonic with Suction
Running Contract Valid Till	10-03-2023
Tender Ref No	KMSCL/EP/T396/862/2020
Tendered Quantity	10
Supplier Name	M/s Healthware Pvt Ltd
GST No	36AABCH2124A1ZE
Installation & Delivery Period	
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Serene Towers 4th Floor - 8-2-623/A Road No. 10 - Banjara Hills Hyderabad-500034	Contact Person	
	Phone	
	Mobile No	
	Email	info@healthwareindia.com,kiran@healthwareindia.com,vipin@healthwareindia.com,rajeevbindra@healthwareindia.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Litho Clast Pneumatic Ultrasonic with Suction <i>Model & Make : FT-128W/EMS ELECTROMEDICAL</i>	3048220.8 Incl.GST :12%	224806.28	3273027.08
2	Ballistic probe - 0.8 mm, length 605 mm	24024 Incl.GST :12%	1771.77	25795.77
3	Ballistic probe - 1 mm, length 605 mm	24024 Incl.GST :12%	1771.77	25795.77
4	Ballistic probe - 1.6 mm, length 453 mm	24024 Incl.GST :12%	1771.77	25795.77
5	Ballistic probe - 2 mm, length 453 mm	24024 Incl.GST :12%	1771.77	25795.77
6	Dual function ultrasound probes: 3.2 mm	13697.6 Incl.GST :12%	1010.2	14707.8
7	Dual function ultrasound probes: 3.6 mm probes	13697.6 Incl.GST :12%	1010.2	14707.8

Item-wise Price Details				
8	Flexible ballistic probe as mentioned in the specifications	16094.4 Incl.GST :12%	1186.96	17281.36
		3187806.4	235100.72	3422907.12

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Litho Clast Pneumatic Ultrasonic with Suction							
Labour	81,648.77	81,648.77	81,648.77	81,648.77	81,648.77	81,648.77	81,648.77
Comprehensive	1,71,462.42	1,71,462.42	1,71,462.42	1,71,462.42	1,71,462.42	1,71,462.42	1,71,462.42

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Litho Clast Pneumatic Ultrasonic with Suction

1. Should have Ballistio and Ultrasound energies to be used simultaneously
2. A single foot pedal to operate both or one energy.
3. Separate hand pieces for Ballistio and ultrasound devices
4. Separate probes and Ultrasound devices
5. Facility to integrate both probes when using both the energies simultaneously to fragment large and hard stones
6. Should be able to use the energies independently also
7. Facility to connect the unit to hospital compressed air supply
8. Should be able to withstand pressures from 3.5 to 5 Bar
9. Facility to connect to the hospital vacuum supply for suction of stone fragments
10. Facility to collect the stone fragments
11. Probes for various applications and scopes
12. The following Ballistic probes should be supplied along with the equipment.
 - a. 0.8 mm, length 605 mm – 2 Nos
 - b. 1 mm, length 605 mm – 4 Nos
 - c. 1.6 mm, length 453 mm – 2 Nos
 - d. 2 mm, length 453 mm – 3 Nos

- For future purchases the rate shall be offered in the BOQ separately (Taken for L1 calculation and the quantity taken for

evaluation will be mentioned in the BOQ)

1. Should have facility for Flexible ballistic probe for use in ureter and renal pelvis through flexible scopes. The rate to be offered in the BOQ separately, if required orders will be issued separately (Not taken for L1 calculation).
2. Should supply Dual function ultrasound probes: 3.2 mm and 3.6 mm probes one each along with the machine. For fragmentation and suction. Probes should have holes in the distal tip to prevent mucosal suction in to the probe.

For future purchases the rate shall be offered in the BOQ separately (Taken for L1 calculation and the quantity taken for evaluation will be mentioned in the BOQ)

1. Conical tip to prevent entry of larger fragments
2. Should supply the guidance adapters for the Ureteroscopes.
3. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid