



Running Contract Details	
Equipment Name	Intubating Video Laryngoscope with Macintosh Blade
Running Contract Valid Till	26-07-2025
Tender Ref No	KMSCL/EP/T489(R)/1272/2023
Tendered Quantity	10
Supplier Name	M/s Lakshmi Surgicals
GST No	32AACFL2603A1ZB
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
2nd Floor 12/307-C1 KSIE Building CIC Freight Station ICTT Road Anavathil Jn. Udyogamandal Eloor Ernakulam - 683501	Contact Person	Mr. Viji Vijayan
	Phone	0484-2988006
	Mobile No	99466 02090
	Email	lakshmisurgicals2005@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Protection Cap, C-MAC Vid Laryng .8403xxx <i>Model &amp; Make : Karl Storz Intubating Video Laryngoscope with Macintosh Blade / Karl Storz GmbH</i>	10682.54 Incl.GST :18%	747.78	11430.32
2	<b>Rolling Stand for C- MAC Monitor</b>	157295.18 Incl.GST :18%	11010.66	168305.84
3	<b>C- MAC Video Laryngoscope MAC#3</b>	463444.8 Incl.GST :12%	34179.05	497623.85
4	<b>C - MAC Â® Video Laryngoscope MAC#0</b>	588114.24 Incl.GST :12%	43373.43	631487.67
5	<b>C-MAC Video Laryngoscope MAC#4</b>	463444.8 Incl.GST :12%	34179.05	497623.85
6	<b>C-MAC Â® Monitor for CMOS Endoscopes, Kit</b>	870658.56 Incl.GST :28%	56184.69	926843.25
7	<b>C- MAC Connecting Cable</b>	82047.76 Incl.GST :18%	5743.34	87791.1

Item-wise Price Details							
8	<b>Protective Bag red for C- MACÂ® System</b>			55766.8 Incl.GST :18%	3903.68	59670.48	
9	<b>MAGILL Forceps ,mod</b>			30726.08 Incl.GST :12%	2266.05	32992.13	
				<b>2722180.76</b>	<b>191587.73</b>	<b>2913768.49</b>	
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Protection Cap, C-MAC Vid Laryng .8403xxx</b>							
<b>Labour</b>	41,300.00	45,430.00	49,973.00	54,970.30	60,467.33	66,514.06	73,165.47
<b>Comprehensive</b>	1,46,000.00	1,60,600.00	1,76,660.00	1,94,326.00	2,13,758.60	2,35,134.46	2,58,647.91

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Protection Cap, C-MAC Vid Laryng .8403xxx**

#### **Equipment: Intubating Video Laryngoscope with Macintosh Blade**

1. Should have Macintosh blades (adult large and extra large 1No each and Paediatric 1 No) with closed European metal finish with integrated camera chip and LED light illumination for obtaining more than 500000 Lux of brightness.
2. Should have Video blades of different sizes and handles constructed of durable, stainless steel and aluminium materials, and drop tested to 48 inches.
3. Should have Screen 7 inch or more in size for display with feature control buttons on the screen with HDMI output for connecting to a big screen. This screen should have option to connect to flexible intubating video endoscope cable as well.
4. Should have High definition and high resolution images features
5. Should have still picture and video recording capability via removable SD card for documentation required
6. Should have Integrated rechargeable battery with a backup of 2 hours.

7. A Soft bag should be supplied to place the monitor and system can also be operated without taking monitor out from the bag
8. IV stand for positioning of the monitor with tray for laryngoscope should be provided
9. Accessories like protection cap, tray for cleaning and sterilization of blades ( at least two blades at a time) should be provided
10. All components should be sterilizable by standard methods.
11. Should provide Magill forceps for foreign body removal and for assisting nasal intubation
12. One imported trolley / box to hang scope as well monitor manufactured by same company should be provided.
13. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.

**Note:** If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid