



Running Contract Details	
Equipment Name	Laser Lithotripter 30 W
Running Contract Valid Till	27-06-2026
Tender Ref No	KMSCL/EP/T516/791A/2023
Tendered Quantity	10
Supplier Name	M/s Allengers Global Healthcare (P) Ltd
GST No	02AAECA8766G1ZP
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Unit 202 2nd Floor - Ansal Classique Tower J-Block Community Centre Rajouri Garden New Delhi - 110027	Contact Person	Sujith Babu
	Phone	
	Mobile No	9946618748
	Email	sujit.babu@allengers.net

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Laser Lithotripter 30 W <i>Model & Make : BLAZE (Holmium YAG Laser)</i>	1672760.32 Incl.GST :12%	123366.07	1796126.39
2	365 micron reusable flexible fiber	26999.84 Incl.GST :12%	1991.24	28991.08
3	200 micron reusable flexible fiber	24999.52 Incl.GST :12%	1843.71	26843.23
4	365 micron stripping and cleaving set	15000.16 Incl.GST :12%	1106.26	16106.42
5	200 micron stripping and cleaving set	15000.16 Incl.GST :12%	1106.26	16106.42
6	Fibre inspection scope	39999.68 Incl.GST :12%	2949.98	42949.66
7	Ceramic Scissor	10000.48 Incl.GST :12%	737.54	10738.02
		1804760.16	133101.06	1937861.22

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Laser Lithotripter 30 W							
Labour	30,000.00	35,000.00	40,000.00	45,000.00	50,000.00	50,000.00	50,000.00
Comprehensive	59,821.00	59,821.00	74,777.00	74,777.00	89,732.00	89,732.00	1,04,688.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Laser Lithotripter 30 W

Equipment: Laser Lithotripter 30W

I. Laser System

1. It should be able to fragment calculi of any size in the bladder, ureter or kidney and any stone fragment.
 2. It should have power output minimum of 30 watts.
 3. It should have repetition rate of 5 -25 Hz.
 4. It should have Energy per Pulse of 0.5 – 3.5 joules
 5. It should have pulse duration of at least 500 microseconds.
 6. It should have Green – aiming beam of 1.0 mw at 532 nm, variable intensity settings.
 7. It should have closed loop self-contained water to air exchanger cooling system.
 8. It should be useable with single phase 230 V AC 50/ 60 HZ.
 9. It should have a shield or self safety mechanism to protect the Laser machine from fiber misalignment or misfire.
 10. List of accessories are given below. The rate for the same shall be offered separately in the BOQ and will be taken for evaluation:
 - a. 365 Micron Reusable , Flexible Fiber
 - b. 200/230 micron reusable, flexible fiber
 - c. 365 Micron Stripping and cleaving (set)
 - d. 200/230 micron stripping and cleaving (set)
 - e. Fibre Inspection Scope (If required for the offered model of equipment, it shall be provided)
 - f. Ceramic Scissors/ diamond cutter
 11. Should supply suitable UPS along with machine and should have minimum back up of 30 minutes
- II. The equipment should have safety certificate from a competent authority CE

issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid.

Note:

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid

2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail.