



Running Contract Details	
Equipment Name	Multiparameter Monitor
Running Contract Valid Till	13-02-2025
Tender Ref No	KMSCL/EP/T468/75/2022
Tendered Quantity	3501
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Multiparameter Monitor <i>Model & Make : K-12 / M/s. Shenzhen Creative Industry Co., Ltd</i>	58128 Incl.GST :12%	4286.94	62414.94
2	Monitor Stand (If needed, order shall be placed separately)	2217.6 Incl.GST :12%	163.55	2381.15
3	Reusable Neonatal SPO2 probe with cable (If needed, order shall be placed separately)	1108.8 Incl.GST :12%	81.77	1190.57
		61454.4	4532.26	65986.66

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Multiparameter Monitor				
Labour	1,500.00	1,500.00	1,500.00	1,500.00
Comprehensive	2,000.00	2,500.00	3,000.00	3,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Multiparameter Monitor

1. Should be able to monitor 6 channel ECG from 5 lead ECG, SPO2, NIBP, Respiration Rate and temperature.
2. Should be portable with carrying handle.
3. Should have touch screen TFT display with at least 12 inches or higher with at least 6 waveforms and numeric display simultaneously.
4. Should have Lithium ion battery with 3 hours battery back up.
5. Should have keys for quick access to main functions.
6. Should have adult, pediatric and neonatal modes.
7. Should provide prominent prioritized audio, visual alarms for high, low heart rate, SpO2, RR, low battery and lethal arrhythmia
8. Recognition with detailed list (more than 15 parameters) in the monitor to set priority and ST Analysis.
9. NIBP can be taken on manual/auto/stat modes.
10. Should provide following accessories:
 - i. Reusable adult 5 lead ECG cable set – 2 nos.
 - ii. Reusable adult and pediatric SPO2 finger probes – 1 each.
 - iii. Adult and pediatric NIBP cuff of different size
 - iv. Rate of Reusable Neonatal SpO2 probe shall be quoted separately which will not be taken for evaluation (If needed order shall be placed separately).
11. Should work on 200-240V AC/50Hz with inbuilt rechargeable battery.
12. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate/ STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.

13. Should have display perfusion index
14. Monitor should have networking facility with bidirectional & bed to bed communication.
15. Monitor stand should be supplied along with the machine as per the following specification, If order placed as per the RC rate.
 - i. Monitor stand -extruded Aluminum, powder coated.
 - ii. Load bearing capacity 20 kgs approximately.
 - iii. Should supply, install with necessary anchor fastners at the site.
16. Equipment performance should not be affected by electromagnetic interferences radiated or conducted through power lines from another device