



Running Contract Details	
Equipment Name	Specular Microscope Type B
Running Contract Valid Till	31-07-2026
Tender Ref No	KMSCL/EP/T526/1015B/2024
Tendered Quantity	4
Supplier Name	M/s My Healthskape Medicals Private Limited
GST No	07AADCM4683K1Z3
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Sharma Industrial Estate Walbhat Road Goregoan East Mumbai 400063	Contact Person	Priyansh vora
	Phone	02226862626
	Mobile No	9820444396
	Email	tenders@myhealthskape.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Specular Microscope Type B</b> <i>Model &amp; Make : EM 4000/ Tomey Corporation</i>	1405027.68 Incl.GST :12%	103620.79	1508648.47
		<b>1405027.68</b>	<b>103620.79</b>	<b>1508648.47</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Specular Microscope Type B</b>							
Labour	15,000.00	15,000.00	15,750.00	15,750.00	16,540.00	16,540.00	17,370.00
Comprehensive	50,000.00	50,000.00	62,500.00	62,500.00	75,000.00	75,000.00	78,750.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Specular Microscope Type B**

#### **â€Equipment: Specular Microscope Type B**

1. Machine should be fully automatic with auto alignment and auto measurement.
2. Analysis should be automatic.
3. Machine should have L -Count, Core and Dark area method.
4. Machine should have integrated Non-Contact Pachymeter.
5. Machine should have 13 measurement areas or more
6. Machine should give count up to 300 cells.
7. Machine should have integrated data base and thermal printer.
8. Machine should be extremely fast.
9. Resolution: Pixels used for picture taking 80 (V) x 180 (H) pixels
10. Capturing scope should be 0.25 x 0.54 mm
11. Photography magnification should be X190
12. Touch Display 10.4'' color LCD
13. Analysis values: CD (cell density), AVG (average cell area), SD (standard deviation of cell area), CV (coefficient of variation of cell area) Cell size (max. + min. cell area)
14. Stroke of moving section X: 88 mm Y: 40 mm Z: 50 mm
15. Stroke of electrical chin rest 70 mm
16. Measuring accuracy Pachymetry +/- 10 µm
17. Data output type USB-Hx2, USB-Dx2, LAN, SD Card (for internal database)
18. Motorized table should be provided along with the machine.
19. Suitable UPS should be provided along with the machine.
20. Machine should be US FDA approved or CE marked.

#### **Note:**

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail