



Running Contract Details	
Equipment Name	Neurosurgical low height Radiolucent surgical table with accessories
Running Contract Valid Till	08-09-2024
Tender Ref No	KMSCL/EP/T411/790/2021
Tendered Quantity	10
Supplier Name	M/s Bet Medical Pvt Ltd
GST No	33AAACB4421H1Z9
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
B-1&B-2 1st Floor Malles Manor No 10 (Old No. 8-A) Periyar Road T Nagar Chennai-600017	Contact Person	Sivakumar
	Phone	
	Mobile No	9995824205
	Email	general@betmedical.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Neurosurgical low height Radiolucent surgical table with accessories <i>Model & Make : TriMax 650NSBM/ BenQ Medical Technology</i>	1829000 Incl.GST :18%	128030	1957030
2	Body strap with 2 clips for fastening side rails - 1 set	8260 Incl.GST :18%	578.2	8838.2
3	Flexible Anesthesia screen with clamps - 1 set	17110 Incl.GST :18%	1197.7	18307.7
4	Goepal knee crutch pair with clamps - 1 set	50740 Incl.GST :18%	3551.8	54291.8
5	Neuro accessory - 1. Connecting bracket for head rest	21240 Incl.GST :18%	1486.8	22726.8
6	Horse shoe shaped rest two piece with connection fixture for flexible positioning	177000 Incl.GST :18%	12390	189390

Item-wise Price Details							
7	Skull clamp with adult and pediatric pins			914500 Incl.GST :18%	64015	978515	
8	Arm positioning device			17700 Incl.GST :18%	1239	18939	
9	Pad for intervertebral disk operation 10 piece design			194700 Incl.GST :18%	13629	208329	
10	Self retaining retractor arm			1180000 Incl.GST :18%	82600	1262600	
11	Pair of shoulder Support			28320 Incl.GST :18%	1982.4	30302.4	
12	Pair of side Support			28320 Incl.GST :18%	1982.4	30302.4	
				4466890	312682.3	4779572.3	
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Neurosurgical low height Radiolucent surgical table with accessories							
Labour	35,000.00	35,000.00	35,000.00	35,000.00	35,000.00	35,000.00	35,000.00
Comprehensive	70,000.00	70,000.00	70,000.00	70,000.00	70,000.00	70,000.00	70,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 3.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Neurosurgical low height Radiolucent surgical table with accessories

Neurosurgical low height Radiolucent surgical table with accessories

1. It should be an electro-hydraulic table, capable of operation by a wired remote control; unit with zero position.

2. It should be able to do all high-end neurosurgical procedures.
3. It should have Hand control and auxiliary control on the column.
4. There should be a manual override with a foot pump for all movements & floor lock and unlock.
5. The operating table should be with easy-rolling castors and the table should have longitudinal shift ≥ 275 mm.
6. Should have an inbuilt battery (rechargeable) with a capacity \geq at least 60 movements.
7. The tabletop should be 4 part table or better.
8. Electro hydraulically operated movements of height, tilting, Backplate and Trendelenburg.
9. Base cover made of glass fiber reinforced plastic – resistant to cracks/breakages and disinfectants or SS304.
10. Tabletop padding should be min 60mm thick.
11. Reverse positioning of patients on a special plate that can be connected at the foot end of the tabletop.
12. Should have wired remote control unit.
13. The table battery should be capable of being charged electrically from line power supply 100-240 VAC, frequency of 50HZ.
14. Width across Backplate ≥ 550 mm, Width across seat plate 540mm or more.
15. Height (Without padding) Min of ≤ 550 mm to Max of ≥ 950 mm.
16. Trendelenburg/reverse trendelenburg : 25-30 degree
17. Lateral tilt, both directions: 15 to 20 degree
18. Backplate up +70/-30 degrees.
19. Max patient weight ≥ 250 kg.
20. Total length of the table should be ≥ 2000 mm and width ≥ 550 mm.
21. Leg plate should be single/split type with padding.
22. It should be excellent ground clearance.
23. It should have high stability while using Drill.
24. Should have Head and leg plates removable.
25. Should be supplied with Arm Board with clamps Qty. 2 no's.
26. Should have radiolucent table-top.
27. Stainless steel side rails on both sides of the tabletop.
28. It should have all certifications for electrical safety.
29. Should have a safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical. **ACCESSORIES (rate to be offered separately in BOQ)**

- a. Body strap with 2 clips for fastening side rails – 1 set.
- b. Flexible Anesthesia screen with clamps – 1 set.
- c. Gopal knee crutch pair with clamps – 1 set.
- d. Connecting bracket for headrest.
- e. Horse shoe-shaped rest two-piece with connection fixture for flexible positioning.
- f. Skull clamp with adult and pediatric pins – Mayfield or Doro (Non-radiolucent skull clamp with adult and pediatric pins) 3 pin skull clamp, Ball pivot adaptor, Base unit, Reusable stainless steel pediatric pin- 3 pins, Reusable stainless steel adult pin – 3 pins)
- g. Arm positioning device – Qty 1 no.
- h. Pad for inter-vertebral disk operation 2 piece design. An accessory for doing spine surgeries. Should be able to do spine surgery with c-arm by the offered accessory
- i. Self-retaining retractor arm used along with Mayfield or Doro skull clamp (Flexible rotary arm 280mm or more-2 pieces, Retractor sterilization tray-1 no, Quick lamp for quick rail-1no., Coupling 2 pieces, Rod-1 piece, Curve arm-1 piece, Sterilization tray-1 No, Blade set 102mm, Black coated 5 pieces-2 sets)
- j. Pair of shoulder support
- k. Pair of Side support.
- l. Table should have integrated continuous X-ray cassette tunnel to slide the cassette to the entire length of the radiolucent tabletop