



Running Contract Details	
Equipment Name	Microdermabrader
Running Contract Valid Till	09-06-2026
Tender Ref No	KMSCL/EP/T522/597/2024
Tendered Quantity	10
Supplier Name	M/s C Cube Advanced Technologies
GST No	29ANCPR4091F1ZS
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
406 3rd floor Malnad Residency 4th Main Hanuma Hills layout Banashankari 3rd stage Bangalore 560061	Contact Person	Vinayaka KS
	Phone	8043021028
	Mobile No	9611111273
	Email	ccubeadvtech@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Microdermabrader <i>Model & Make : Tower Model/ C Cube</i>	91840 Incl.GST :12%	6773.2	98613.2
2	Micro needle tips	531 Incl.GST :18%	37.17	568.17
		92371	6810.37	99181.37

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Microdermabrader				
Labour	5,000.00	5,000.00	5,000.00	
Comprehensive	10,000.00	10,000.00	10,000.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Microdermabrader

1. The unit should be a tower model.
2. The unit should have strong base with 4 wheels of size at least 2 inch (2 with brake and 2 without brake)
3. The unit should deliver at least 600 mmHg vacuum
4. The unit should have auto cut off facility.
5. The system should have provision for setting the time and a countdown timer to enable auto cut off feature should also be available when the set time is elapsed.
6. The unit should have cumulative hour meter with display
7. The vacuum should be adjustable in fine variations using rotary turned knob and should have a display
8. The unit should have vacuum system of 0.25 HP oil free pump
9. The unit should have security on/off locking system
10. The unit should work with 230V, 50Hz for protection against leakage current
11. The unit should have solenoid valves to ensure proper switch over from crystal to diamond tip
12. The unit should be able to operate with crystal powder tips and diamond tips
13. Three metal tips to be provided for crystal powder operation along with the machine free of cost and one metal tip for diamond powder operation
14. The unit should have burst mode for crystal powder operation using foot switch
15. The unit should have facility to take out used powder
16. Should have non return valve to ensure unidirectional crystal powder flow.
17. The unit should have tip holder to keep the accessories
18. Should supply 5 kg Aluminium oxide crystal powder free of cost
19. Should have micro needle skin nursing system to treat acne and scar removal
20. Should be able to provide 2 Bar pressure compressed air to clean crystal powder after treatment
21. The unit should have micro needling system
22. Should have 415nm laser pen to treat inflammation in acne
23. Should have galvanic hand piece for skin rejuvenation
24. The unit should be ISI / CE certified and the copy should be submitted

Note:

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail