



Running Contract Details	
Equipment Name	Transport Ventilator
Running Contract Valid Till	25-08-2024
Tender Ref No	KMSCL/EP/T446/1319/2022
Tendered Quantity	20
Supplier Name	M/s C.S.Associates
GST No	33AAEPC2663B1ZH
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Robertson Lane Mandaveli Chennai	Contact Person	Thilakraj C
	Phone	04442075974
	Mobile No	9600131594
	Email	contact@csassociates.info

Item-wise Price Details							
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	Transport Ventilator <i>Model & Make : TV-100/Bio-Med Devices</i>	1366400 Incl.GST :12%	100772	1467172			
		1366400	100772	1467172			
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Transport Ventilator							
Labour	30,000.00	30,000.00	30,000.00	30,000.00	30,000.00	30,000.00	30,000.00
Comprehensive	60,000.00	60,000.00	60,000.00	60,000.00	60,000.00	60,000.00	60,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Transport Ventilator

1. Transport ventilator should be portable, light weight.
2. It should be Time cycled pressure limited dedicated Neonatal ventilator.
3. Inspiration time: 0.3 to 2 sec.
4. Expiration time: 1 to 10 sec.
5. PIP should be 10 cm H₂O to 40 cm H₂O.
6. PEEP should be 1 cmH₂O to 15 cmH₂O.
7. Tidal volume delivered should be between 5 ml - 100 ml
8. Modes - assist control, SIMV Pressure control, CPAP mode
9. Spares:
 - i. 10 L oxygen cylinder with flowmeter and pressure indicator : 2 Nos
 - ii. Reusable breathing circuits for transport ventilator : 4 Nos
 - iii. Reusable nasal prongs for CPAP : 4 Nos
 - iv. 12 V DC mini compressor for transport ventilator : 1 Nos
10. Battery backup should be four hours
11. FiO₂ should be 21 to 100%