



Running Contract Details	
Equipment Name	Iron Cot
Running Contract Valid Till	29-05-2025
Tender Ref No	KMSCL/EP/T489/364/2022
Tendered Quantity	2000
Supplier Name	M/s Midmark (India) Private Limited
GST No	24AACCM0219N1ZM
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	0 Years

Supplier`s Details		
Address	Contact Details	
A 1 3 GIDC UMBERGAON Valsad Gujarat - 396171	Contact Person	Mr. Renjith Lal
	Phone	+91 260 2562 406
	Mobile No	9605680008
	Email	pkadam@midmark.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Iron Cot <i>Model & Make : BX1000 Basic Simple Bed / M/s. Midmark (India) Private Limited</i>	12378.2 Incl.GST :18%	866.47	13244.67
		12378.2	866.47	13244.67

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Iron Cot

1. Overall dimension shall be 2000mm x 900mm x 600mm (LxWxH) tolerance ± 20 mm
2. Frame dimension shall be 1980mm x 900mm (LxW) tolerance ± 20 mm
3. Mattress dimension shall be 1980mmx900mm tolerance ± 20 mm
4. The top sheet shall be 18 gauge(1.2mm) thickness CRCA sheet double bend of height minimum reinforced with channels of size minimum 75mm x 30mm flange welded inside the lengthwise frame bends.
5. The bed frame should be provided widthwise 3 nos of angle stiffener of size minimum 35x35x2mm and 2 nos angle support size 50x35x2mm one at each head and leg.
6. The leg frames of the cot shall be made of MS tube of diameter 38 mm thick with horizontal bottom tube of diameter 25x1.2 thickness and vertical filling tube of diameter 16mm x 1.2mm thickness.
7. The horizontal cross tube should be welded at a height of 270mm from the bottom.
8. The leg frame at the head end should have a height of 1075mm and leg side should be at least 820mm.
9. Pretreated and powder coated finish
10. Angular bracket of 3mm thickness shall be used for fixing the bed frame
11. Sockets should be provided for Mosquito net poles 20mmdiameter and 26mm length
12. The mattress shall be 4” thick, PU foam of 40 density covered with PVC Rexine.
13. Should have provisions to connect IV poles on all four corners of the cot.
14. Should be provided with leg shoes made of PVC bush of height at least 40mm, outside diameter should be 40mm and thickness at least 4 mm.
15. All bolts and nuts used shall be electroplated.
16. The manufacturer should compliant with ISO 9001, 14001, 13485 & OHSAS 18001/OHSAS 45001

Note: If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid