



Running Contract Details

| | |
|--------------------------------|---|
| Equipment Name | Composite light Cure Unit |
| Running Contract Valid Till | 08-05-2026 |
| Tender Ref No | KMSCL/EP/T517(R)/400D/2023 |
| Tendered Quantity | 100 |
| Supplier Name | M/s X-Ray Medico Surgical Engineering Ancillary |
| GST No | 33AAAFX0226H1ZM |
| Installation & Delivery Period | 8 Week(s) |
| Up-time / PM vist | 95% & 2 Visits per year |
| Warranty period | 1 Years |

Supplier`s Details

| Address | Contact Details | |
|--|-----------------|--------------------|
| Old No.45 New No.64 Third Main Road Kasthurba Nagar Adyar Chennai-600020 | Contact Person | N. Ramanathan |
| | Phone | 044 24410074 |
| | Mobile No | |
| | Email | xmse6445@gmail.com |

Item-wise Price Details

| # | Item Details | Unit Rate (Incl.all taxes & charges) | Service Charges (Through KMSCL) | Grand Total |
|---|---|---|------------------------------------|--------------|
| 1 | Composite light Cure Unit <i>Model & Make : WOOD PECKER(LED-D)/ X-RAY MEDICO SURGICAL ENGINEERING ANCILLARY</i> | 35400 Incl.GST :18% | 2478 | 37878 |
| | | 35400 | 2478 | 37878 |

Annual / Comprehensive Maintenance Charges (Exl.Tax)

| Rate | 2 ND Year | 3 rd Year | 4 th Year |
|----------------------------------|----------------------|----------------------|----------------------|
| Composite light Cure Unit | | | |
| Labour | 1,000.00 | 1,000.00 | 1,000.00 |
| Comprehsi ve | 1,000.00 | 1,000.00 | 1,000.00 |

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Composite light Cure Unit

- 1.Should have a LED lamp and should have a calibrated automatic timer for selection of cure time from 10 to 60 seconds. The cost of replacement of LED has to be included in the warranty as well as CAMC offered.**
- 2.It should be a Composite Visible Light Cure Unit**
- 3.Should have a light intensity range of 850-1000 mw/cm square or more.**
- 4.The hand piece should have on/off trigger and audible tones in every 10 sec.**
- 5.Should supplied with autoclavable fiber optic probe with full 360° rotation & it should be a 60 degree angleprobe**
- 6.Should be separate unit not attached to Dental unit**
- 7.Should provide an Eye protection shield**

Note:

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail