



Running Contract Details	
Equipment Name	Ventilator CPAP (Non-Invasive)
Running Contract Valid Till	27-04-2024
Tender Ref No	KMSCL/EP/T416(R)/136A/2021
Tendered Quantity	500
Supplier Name	M/s ResMed India Pvt Ltd
GST No	07AADCR5533K1Z7
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
205 IInd Floor Southern Pard D-2 District Centre Saket New Delhi - 110017	Contact Person	Mr. Vinay Rattan
	Phone	
	Mobile No	9560299926
	Email	enquiries@resmed.co.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Ventilator CPAP (Non-Invasive)</b> <i>Model &amp; Make : Airsense</i>	48473.6 Incl.GST :12%	3574.93	52048.53
		<b>48473.6</b>	<b>3574.93</b>	<b>52048.53</b>
Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year	
<b>Ventilator CPAP (Non-Invasive)</b>				
Labour	857.00	857.00	857.00	
Comprehensi ve	1,714.00	1,714.00	1,714.00	

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Ventilator CPAP (Non-Invasive)**

#### **Equipment Name: Ventilator CPAP(Non Invasive)**

1. Should have Manual and Auto CPAP mode with advanced sensor technology for tracking respiratory status of the patient.
2. It should be suited for therapy of OSA / Obesity hypoventilation therapy.
3. Should have inspiration trigger for auto start.
4. Should have auto stop option when put-off mask.
5. Should have alarms for power off, disconnection.
6. Should have facility for sensitivity setting.
7. Should have leakage compensation facility.
8. Should have integrated DC Powered humidifier or equivalent with anti- counter current water trap.
9. Should have embedded memory storage for full raw data and facility for longer storage.
10. Should be supplied with standard spares and accessories
11. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.
12. Should have ramp settings for CPAP Pressure.
13. Should have mask size check feature
14. Should have CPAP Pressure (4-20) cm of H<sub>2</sub>O
15. Should have display of at least 2 inches
16. Should have inbuilt battery backup of 1 hour