



Running Contract Details	
Equipment Name	Sphygmomanometer Stand model
Running Contract Valid Till	23-07-2026
Tender Ref No	KMSCL/EP/T522/113/2024
Tendered Quantity	75
Supplier Name	M/s Trivandrum Surgicals and Lab Supplies
GST No	32ERJPS6054N1Z5
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
T.C.2/1094(7) SREELAKSHMI PAZHAYA ROAD MEDICAL COLLEGE P.O TVM.695015	Contact Person	GEETHA . S
	Phone	0471 - 2555040
	Mobile No	09447340147
	Email	trivandrumsurgicals@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Sphygmomanometer Stand model <i>Model & Make : DIAMOND BPDG 534/INDUSTRIAL ELECTRONIC & ALLIEDPRODUCTS</i>	8499.99 Incl.GST :12%	626.87	9126.86
		8499.99	626.87	9126.86

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Sphygmomanometer Stand model

Equipment Name: Sphygmomanometer – Stand Model

1. Should be portable digital type, stand model.
2. Should have LED/LCD indication for pressure
3. Should have a measuring range from 0 to 300 mmHg.
4. Should be provided with latex free adult arm cuffs of size medium & large and pediatric cuff.
5. The manometer scale markings and graduations should be permanent and clearly visible
6. Should have measurement accuracy of pressure +3mmHg.
7. Should have battery life of approx .1500 measurements
8. Should have adjustable display column for easier analysis of result.
9. The housing case should be robust design ABS or metal. It should have a cuff holder.
10. The latex free tubes used should have an internal diameter of 3 ± 0.5 mm and the external diameter should not be less than 8mm..
11. The tubes should be fitted with male and female leur connectors.
12. The inflating bulb should be soft and should not have any joints or ridges.
13. Should be mounted on good quality wheels.
14. The stand body shall be made of mild steel and powder coated.