



Running Contract Details	
Equipment Name	Analytical Balance with Gram Sensitivity
Running Contract Valid Till	03-08-2023
Tender Ref No	KMSCL/EP/T391/280B/2021
Tendered Quantity	20
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Analytical Balance with Gram Sensitivity <i>Model & Make : CY 124 C/ Aczet</i>	76464 Incl.GST :18%	5352.48	81816.48
		76464	5352.48	81816.48
Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	4 th Year	5 th Year	6 th Year	
	Analytical Balance with Gram Sensitivity			
Labour	900.00	900.00	900.00	
Comprehensive	1,500.00	2,000.00	2,500.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 3.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Analytical Balance with Gram Sensitivity

Analytical Balance With Gram Sensitivity

1	Read out	0.1
2	Weighting range	120 g
3	Reproducibility	0.2mg
4	Linearity	0.3mg
5	Calibration	Internal
6	Display	LCD, Digit height 14mm
7	Weighting plate	91mm diameter
8	Weighting cell	Single cell.
9	Stabilization time	Steady weight values within 4 sec under laboratory conditions
10	Construction	Shock proof
11	Certifications	CE