



**Running Contract Details**

Equipment Name	Defibrillator with Cardiac Monitor
Running Contract Valid Till	11-04-2026
Tender Ref No	KMSCL/EP/T522/27/2024
Tendered Quantity	200
Supplier Name	M/s Schiller Healthcare India Pvt. Ltd
GST No	34AADCS5091B1ZN
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

**Supplier`s Details**

Address	Contact Details	
Schiller Healthcare India Pvt Ltd 37/1013 - 2nd Floor - Unidec House Fathima Church Road Elamkulam - 682 020	Contact Person	Mr.Stijo Mathew, Mr. P.S Vidhyashankar
	Phone	04842203264
	Mobile No	09846706787, 09383620520
	Email	alok@schillerindia.com, Mail@schillerindia.com, shinoj@schillerindia.com, kishorekumar@schillerindia.com, stijo@schillerindia.com, vidyashankar@schillerindia.com, mail@schillerindia.com, kishorekumar@schillerindia.com

**Item-wise Price Details**

#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Defibrillator with Cardiac Monitor</b> <i>Model &amp; Make : Defigard 400/Schiller Healthcare India Pvt. Ltd.,</i>	188440 Incl.GST :12%	13897.45	202337.45
		<b>188440</b>	<b>13897.45</b>	<b>202337.45</b>

**Annual / Comprehensive Maintenance Charges (Exl.Tax)**

Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Defibrillator with Cardiac Monitor</b>							
Labour	3,500.00	3,500.00	3,500.00	3,500.00	3,500.00	3,500.00	3,500.00
Comprehensive	9,500.00	9,500.00	9,500.00	9,500.00	9,500.00	9,500.00	9,500.00

## **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Defibrillator with Cardiac Monitor**

1. Biphasic, Manual and AED with voice prompt, compact and light weight
2. Energy selection 5J to 200J insteps.
3. Momentary energy selection access on front panel.
4. Should have adult and pediatric paddles integrated on same handle.
5. Momentary charge key on front panel and on the apex hand.
6. Monitor should display selected and delivered energy
7. Should have disarm facility.
8. Energy should be delivered within 60ms after the detected R wave in synchronization mode.
9. Charging time maximum 6 sec for 200J.
10. Should have battery backup for 50discharges of 200J.
11. Should have ECG inputs through paddles or 3 lead cables.
12. Should have display for selected ECG input source(I, II, III, paddles)
13. Lead off message should appear with alert tone.
14. Amplitude gain of ECG waveform should be adjustable
15. Should have colour display for heart rate of size 6 inches or more.
16. Should have alarm for high and low HR.
17. Should have an inbuilt thermal recorder.
18. Should have enable/disable option for printer.
19. Should supply 2 bottle of jelly, 12 roll of thermal paper.
20. Should supply three pairs of AED pads and the prices of AED Pads should be quoted separately which will not be taken for evaluation.
21. Should operate on mains 230V, 50Hz
22. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid

### **Note:**

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail