

(A Government of Kerala Undertaking) Thycaud P.O, Thiruvananthapuram - 14, Kerala. Tel: 0471 - 2945600, 2337353, Fax: 0471 - 2945647 Email :engg_ep@kmscl.kerala.gov.in CIN: U24233KL200TSGC021616, PAN : AADCK4029M, GSTIN : 32AADCK4029M1ZK

Running Contract Details						
Equipment Name	Semi Auto Biochemistry Analyser					
Running Contract Valid Till	22-02-2025					
Tender Ref No	KMSCL/EP/T467/108/2022					
Tendered Quantity	300					
Supplier Name	M/s Meril Diagnostics Private Limited					
GST No	24AAHCM1404G1ZZ					
Installation & Delivery Period	8 Week(s)					
Up-time / PM vist	95% & 2 Visits per year					
Warranty period	2 Years					

Supplier`s Details								
Address	Contact Details							
601 - Midas Tower - Sahar Plaza	Contact Person	Rajendra Jagtap						
Andheri(E) Mumbai 400059	Phone							
	Mobile No	8422802267, 8422940405						
	Email	prakash.teckchandani@merillife.com, gigi.jo y@merillife.com,tender.merildiagno@merill ife.com						

Item-wise Price Details										
#		Item Details			Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)		Grand Total		
1	Semi Auto Biochemistry Analyser Model & Make : Merillizer CliniQuant Micro / M/s. Meril Diagnostics Pvt Ltd			49324 Incl.GST :18%	3452.68		52776.68			
2	Additional Filter 1 (Other than 6 inbuilt free filters)			1180 Incl.GST :18%	82.6		1262.6			
3	Additional Filter 2 (Other than 6 inbuilt free filters)			1180 Incl.GST :18%	82.6		1262.6			
					51684	3617.88		55301.88		
Annual / Comprehensive Maintenance Charges (Exl.Tax)										
Rate	:	3 rd Year		4 th Ye	r		5 th Year			
Semi Auto Biochemistry Analyser										
Labou	Labour 300.00		300.00		300.00					
Compr ve	Comprehensi 1,400.00		1,400.00		1,400.00					

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.

3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.

4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Semi Auto Biochemistry Analyser

- 1. Should be microprocessor controlled general purpose bi-chromatic photometer system with at least 6 filters ranging from 340 to 630nm.
- 2. Temperature 37 self monitoring built-in incubation systems for temperature controlled absorbance reading.
- 3. Light source : Tungsten/ halogen or higher grade with one additional bulb.
- 4. Should have end point, kinetic and two point kinetic measurement modes.
- 5. Should have flow cell measuring device.
- 6. Should have inbuilt printer or External Printer to be supplied.
- 7. Should have a measurement range from 0.001 to 2.300Abs
- 8. Should have facility for reading results on LCD display.
- 9. Should have quality control two control/test QC survey of at least 30 points, Levy Jenny plot.
- 10. Should have a filter half bandwidth of 10nm or lesser.
- 11. Should have a test programme memory of 50 or more.
- 12. Should be provided with sample carry over prevention facility.
- 13. Aspiration should be based on Bellow/Peristaltic Pump
- 14. Should provide 500 ml of reagents for urea, S. creatine, S. bilirubin, sugar, cholesterol, and Quality control 5ml one each for normal, abnormal.
- 15. Should be supplied with UPS of suitable capacity with a minimum back of 30 minutes.
- 16. Should be provided with calibration certificate issued by the manufacturer at the time of installation and periodic calibration must be done and calibration certificate to be issued during warranty / AMC/ CAMC period.
- 17. Should have safety certificate from a competent authority CE-IVD issued by a notified body registered in the European commission / FDA (US). Copy of the certificate shall be produced along with the technical bid.
- 18. Should have inbuilt / external incubator with 6 / 8 wells along with the machine