

Running Contract Details	
Equipment Name	Monocular Microscope with inbuilt Light Source
Running Contract Valid Till	31-07-2025
Tender Ref No	KMSCL/EP/T486/192B/2023
Tendered Quantity	100
Supplier Name	M/s Radical Scientific Equipments Pvt Ltd
GST No	06AACCR8985N1ZI
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
9th Milestone Ambala-Jagadhri Road Khudda Kalan Ambala Cantt-133104 (Haryana)	Contact Person	Mr. Vikram Singh Chauhan
	Phone	
	Mobile No	89030997827
	Email	tenders@radicalscientific.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Monocular Microscope with inbuilt Light Source <i>Model & Make : RXL-4/Radical Scientific Equipments Pvt. Ltd.</i>	15340 Incl.GST :18%	1073.8	16413.8
		15340	1073.8	16413.8
Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Monocular Microscope with inbuilt Light Source				
Labour	180.00	180.00	180.00	
Comprehensive	210.00	210.00	210.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Monocular Microscope with inbuilt Light Source

Equipment: Monocular Microscope with Inbuilt Light Source

1. Should have quadruple nose piece and DNI plan achromatic objectives 4x, 10x, 40x (SL), 100x OIL (SL) anti-fungus coating
2. Should have LED illumination with rechargeable battery cells to provide battery backup in power failure
3. Input should be 220-240 VAC
4. Should have steady focus control with convenient location and ball bearing.
5. Should have monocular tube 45 degree inclined and 360 degree rotatable
6. Should have integrated wide field eyepiece-10x
7. Should have Anti-fungal coated optics
8. Should have quadruple nose piece with ball bearings
9. Mechanical stage of dimension: 135 x 125 mm with coaxial control with ball bearing. Graduated scale.
10. Should have coaxial coarse and fine focusing system with ball bearings
11. Illumination: LED illumination with in-built rechargeable battery and (5-6) hour's battery backup in case of electricity failure.
Life of the LED source should be at least 1500 hours
12. Should have stainless steel holder for slide with 75 x 50 mm movement
13. Should have Abbe condenser, NA 125 aspherical lens
14. Should have safety certificate from a competent authority CE issued by a notified body registered in European commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid

Note: If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid