



Running Contract Details	
Equipment Name	Vertical Autoclave - Small(Portable)
Running Contract Valid Till	30-01-2025
Tender Ref No	KMSCL/EP/T467/9E/2022
Tendered Quantity	1500
Supplier Name	M/s New India Surgicals and Chemicals
GST No	32AAGFN8445E1ZD
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
30/1212IV Prasanth Buildings Near Medical College Hospital Calicut-673008	Contact Person	Prasanth N
	Phone	0495-2354188
	Mobile No	9447054188
	Email	nisc2000@rediffmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Vertical Autoclave - Small(Portable)</b> <i>Model &amp; Make : KAUC-PS / Kadavil Electro Mechanical Industries</i>	11682 Incl.GST :18%	817.74	12499.74
		<b>11682</b>	<b>817.74</b>	<b>12499.74</b>

### Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Vertical Autoclave - Small(Portable)**

1. Chamber volume should be approximately 300mm diameter and 500mm depth.
2. Should have stainless steel construction
3. Should have pressure gauges for measuring chamber pressure.
4. Should have safety valve and weight valve.
5. Should have 121°C working temperature at 15 psi pressure.
6. Should be operated in mains 220 – 240 V ac 50 Hz input power supply.