

Running Contract Details	
Equipment Name	T- Piece resuscitator
Running Contract Valid Till	03-10-2023
Tender Ref No	KMSCL/EP/T403(R)/1216/2021
Tendered Quantity	100
Supplier Name	M/s Nice Neotech Medical Systems Pvt Ltd
GST No	33AABCN4311M1ZO
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
No.85 Krishna Industrial Estate Vannagaram Mettukuppam Chennai 600 095 Tamil Nadu	Contact Person	K.Premkumar
	Phone	
	Mobile No	9543455323
	Email	sales@niceneotech.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>T- Piece resuscitator</b> <i>Model &amp; Make : Nice 5020/ nice Neotech</i>	57568 Incl.GST :12%	4245.64	61813.64
		<b>57568</b>	<b>4245.64</b>	<b>61813.64</b>

### Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 3.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :T- Piece resuscitator**

#### **T-Piece Resuscitator**

1. Should be manually operated, gas-powered resuscitator with consistent Peak Inspiratory Pressure(PIP) and Positive End Expiratory Pressure (PEEP) to infants.
2. Should accept and deliver oxygen concentrations from 21% to 100%.
3. T-piece circuits should be connected to infant resuscitation masks or endotracheal tubes.
4. The gas flow outlet should allow easy and secure connection of the T-piece circuit and humidified T-piece circuit.
5. Manometer range : -10 to 80 cmH20;  $\pm$  2% accuracy
6. Maximum pressure relief : 8L/min 5 to 70 cmH20
7. Peak Inspiratory Pressure : 5L/min 2 to 70 cmH20

8L/min 3 to 72 cmH20

10L/min 4 to 73 cmH20

15L/min 8 to 75 cmH20

8. Input gas flow range: 5L/min (minimum) to 15L/min (maximum)

9. Should be supplied with the following accessory

10. Gas supply line – 3 Nos

11. Should be supplied with the following consumables

- a . Disposable infant resuscitation T-piece circuit – 10 Nos
- b . Disposable infant resuscitation T-piece mask – 10 Nos
- c . Disposable humidifier T-piece circuit with heated wire – 10 Nos

- 12 . Should have a safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ STQC CB Certificate/ STQC S Certificate or a valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid