



**Running Contract Details**

Equipment Name	A-scan
Running Contract Valid Till	04-02-2026
Tender Ref No	KMSCL/EP/T499/6/2023
Tendered Quantity	30
Supplier Name	M/s Appasamy Associates PVT. LTD
GST No	33AASCA3660R1Z0
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

**Supplier`s Details**

Address	Contact Details	
# 20 - SBI Officers Colony 1st Street Arumbakkam Chennai-600106	Contact Person	B Radhakrishnan
	Phone	04430101400, 30101401
	Mobile No	09349211170
	Email	info@appasamy.com, appasamy.thrissur@gmail.com

**Item-wise Price Details**

#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>A-scan</b> <i>Model &amp; Make : Appascan Max/ Appasamy Associates Pvt Ltd.</i>	252000 Incl.GST :12%	18585	270585
		<b>252000</b>	<b>18585</b>	<b>270585</b>

**Annual / Comprehensive Maintenance Charges (Exl.Tax)**

Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>A-scan</b>							
Labour	6,700.00	6,700.00	6,700.00	6,700.00	6,700.00	6,700.00	6,700.00
Comprehensive	13,500.00	13,500.00	15,700.00	15,700.00	18,000.00	18,000.00	20,200.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :A-scan**

1. Should have easy readable LCD display
2. Should have adjustable gain control
3. Should have a probe frequency of 10Mhz or more with internal fixation light
4. Should use immersion & contact measurement technique
5. Should have automatic, manual, PMMA, Silicone, Acrylic modes.
6. Should provide IOL formulas such as Holladay II, Hoffer Q, SRK/T, Haigis etc and should have accepted post refractive formulas like shammas or equivalent.
7. Should have a measurement range of 14 to 40mm or more
8. Should measure Anterior Chamber Depth, VCD, Lens thickness and axial length
9. Should have inbuilt thermal printer / external printer through USB/LAN
10. Should provide footswitch / automode, immersion cup, carry case and probe holder.
11. Should have LAN/USB/ RS232 Interface port connectivity
12. Should operate from 200 to 240V / 50 Hz power supply
13. Should have safety certificate from a competent authority CE issued by a notified body registered in the European commission / FDA (US)/ /IC-MED / STQC CB Certificate/ STQC S Certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/ test report shall be produced along with the technical bid.

#### **Note:**

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail