



Running Contract Details	
Equipment Name	Lead Apron
Running Contract Valid Till	08-01-2026
Tender Ref No	KMSCL/EP/T512/71/2023
Tendered Quantity	250
Supplier Name	M/s C Cube Advanced Technologies
GST No	29ANCPR4091F1ZS
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
406 3rd floor Malnad Residency 4th Main Hanuma Hills layout Banashankari 3rd stage Bangalore 560061	Contact Person	Vinayaka KS
	Phone	8043021028
	Mobile No	9611111273
	Email	ccubeadvtech@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Lead Apron <i>Model & Make : AT-100 /AMRAD MEDICAL EQUIPMENTS</i>	10620 Incl.GST :18%	743.4	11363.4
		10620	743.4	11363.4

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Lead Apron

1. Should be AERB approved.
2. Should be light weight 0.5mm lead equivalent.
3. Should be hook and loop type (Velcro).
4. Should be supplied along with thyroid guard

Note:

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail