



Running Contract Details	
Equipment Name	OT Table Electro Hydraulic
Running Contract Valid Till	02-12-2023
Tender Ref No	KMSCL/EP/T413/197A/2021
Tendered Quantity	100
Supplier Name	M/s Surya International Enterprise Private Limited
GST No	21ABFCS9529K1Z0
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
A - 24 HIG Duplex Phase 2 Ananta Vihar Pokhariput Odisha - 751020	Contact Person	Kamlesh Kumar Singh
	Phone	
	Mobile No	(+91) 9439525334
	Email	nbi@groupsurya.co.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	OT Table Electro Hydraulic <i>Model & Make : Anya Surgical & Healthcare Enterprises</i>	561320.1 Incl.GST :18%	39292.41	600612.51
2	Horse Shoe Shaped Head rest(ss Grade 304)	94400 Incl.GST :18%	6608	101008
3	Adapter to fix three pin skull clamp(ss grade 304)	59000 Incl.GST :18%	4130	63130
4	Hand operating table with height adjustable floor support	64900 Incl.GST :18%	4543	69443
5	Raised arm Board with lock	7670 Incl.GST :18%	536.9	8206.9
6	Gel Body Roll for adult(Size-510 mmL X 155mmW X 125mmH)	21240 Incl.GST :18%	1486.8	22726.8
7	Electrocautery earthing sheet	41300 Incl.GST :18%	2891	44191

Item-wise Price Details							
				849830.1	59488.11	909318.21	
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
OT Table Electro Hydraulic							
Labour	37,000.00	37,000.00	37,000.00	37,000.00	37,000.00	37,000.00	37,000.00
Comprehensive	20,000.00	20,000.00	20,000.00	20,000.00	20,000.00	20,000.00	20,000.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 3.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :OT Table Electro Hydraulic

OT Table Electro-Hydraulic

1. The table should have a minimum of 4 sections ie. head section, leg section, a seat section, and backplate section.
2. The table should be electro-hydraulically operated having the following hand switch operated electro-hydraulic functions (all the dimensions will have a permitted deviation of +/- 10 %)

Sl No	Description	Range
i.	Up / Down	680-1000mm
ii.	Trendelenberg& Reverse	30deg
iii.	Side Tilting (Lateral)	30deg
iv.	Back Plate (Sitting Position)	-40 to +80
v.	Top Slide	300mm
vi.	Braking (by hand switch)	

3. All the above functions must have a manual override Override System present on the column for up/down, Trendelenburg /reverse Trendelenburg and tilt, Floor locking.

4. In addition to the above-hand switch operated functions, the table must have the following manual functions.

	Description	Range
i.	Head Section Tilting	30deg Up / 90deg Down
ii.	Split Leg Plate manual movement	+10 to -90

5. The tabletop must be made of durable radiolucent bakelite material capable of withstanding exposure to frequent C-Arm imaging, without diminishing the image clarity.

6. The table should be supplied with the following accessories.

- i. Mattress for the complete tabletop in sections - 1 set
- ii. A pair of arm boards with pad and fixing clamp - 1 set
- iii. A pair of padded shoulder support with clamps (SS grade 304) – 1
- iv. A pair of padded lateral support with clamps (SS grade 304) – 1
- v. Anesthetic screen frame with clamp (SS grade 304) – 1
- vi. Patient restraint strap – 1
- vii. leg crutches with side rail locks – 1 pair .

7. The base cover, lifting column cover and side rails should be made of stainless steel grade SS grade 304

8. Should have the enhanced weight bearing casters fitted with ball bearing.

9 . The table should have a heavy and sturdy base and be compact to provide adequate foot room for the operating team.

10. The weight-bearing capacity of the table shall be at least 175kg.

11. Should have a safety certificate from a competent authority CE / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid.

12. battery backup shall be available for minimum 50 functions.

13. Should have following accessories and the rate for the same should be offered separately.

1. Horseshoe-shaped headrest (SS grade 304) and adapter to fix three-pin skull clamp (SS grade 304) - 1set.
2. Hand Operating Table with height adjustable floor support - 1 set.
3. Raised arm Board with lock - 1 No.
4. Gel Body Roll for adult - 1 No (Size - 510mmL X 155mmW X 125mmH).
5. Electrocautery earthing sheet – 1 No