



Running Contract Details	
Equipment Name	MRI Compatible Ventilator
Running Contract Valid Till	29-09-2024
Tender Ref No	KMSCL/EP/T446/618/2022
Tendered Quantity	10
Supplier Name	M/s Techsure Medical Devices Pvt. Ltd.
GST No	32AAECM4888R1ZO
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
16/1080 9 B 2 BCG TOWER KAKKANAD ERNAKULAM-682037	Contact Person	SUNILKUMAR JAYADEV
	Phone	484850073
	Mobile No	9388880678
	Email	mail@techsureindia.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	MRI Compatible Ventilator <i>Model & Make : HAMILTON-MRI/Hamilton Medical AG</i>	3427200 Incl.GST :12%	252756	3679956
		3427200	252756	3679956

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
MRI Compatible Ventilator							
Labour	1,22,400.00	1,28,520.00	1,34,946.00	1,41,693.00	1,48,778.00	1,56,217.00	1,64,027.00
Comprehensive	3,06,000.00	3,36,600.00	3,70,260.00	4,07,260.00	4,48,014.00	4,92,816.00	5,42,097.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :MRI Compatible Ventilator

1. Ventilator should be MR compatible with 1.5 and 3 Tesla
2. Should have integrated Tesla Spy magnetic field navigator
3. Machine should be portable with independence from compressed air
4. Should have paediatric, neonatal and adult modes of ventilation
5. Should provide minimum 8 hrs battery operating time
6. Should also provide Non- Invasive ventilation and integrated high flow oxygen therapy
7. Should have auto lock facility for trolley when handles are released
8. With should have facility to automatically switch between gas cylinder and central oxygen supply
9. Should have the following modes; pressure control ventilation, volume control ventilation, pressure support ventilation, volume support ventilation, C PAP-PS
10. Should have configurable loops and trends
11. Integrated nebulizer
12. Well visible 360 degree lamp alarm
13. Should provide humidifier