



Running Contract Details	
Equipment Name	Electric Dermatome
Running Contract Valid Till	16-03-2025
Tender Ref No	KMSCL/EP/T472/874/2022
Tendered Quantity	5
Supplier Name	M/s Praxis Health Technologies
GST No	32AAPFP1540D1ZM
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Puthiyapalam Tali Kozhikode Chalappuram P.O	Contact Person	Anoop M
	Phone	4952700425
	Mobile No	9747275757
	Email	praxis.health2013@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Electric Dermatome <i>Model & Make : E100D/ G.R.Bioure Surgical System Pvt Ltd</i>	601692 Incl.GST :12%	44374.79	646066.79
		601692	44374.79	646066.79

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Electric Dermatome							
Labour	10,750.00	10,750.00	10,750.00	10,750.00	10,750.00	10,750.00	10,750.00
Comprehensive	26,861.00	28,204.00	29,614.00	31,095.00	32,650.00	34,282.00	35,997.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).

2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Electric Dermatome

1. It is a skin grafting instrument to provide variable graft thickness and width capabilities
2. Should have an electrically-powered surgical skin grafting instrument.
3. Should have thickness control adjustment ranges from 0 to 0.030 in. (0.75 mm) in 0.002 in. (0.050 mm) increments.
4. Should have Individual graft widths of 1 in., 1.5 in., 2 in., 3 in. and 4 in. (2.5 cm, 3.8 cm, 5.1 cm, 7.6 cm, 10.2 cm) are obtained with five width plates.
5. Should have two stainless steel machine screws secure the plates to the underside of the instrument.
6. The plates are easily fastened and removed with the screwdriver provided.
7. The dermatome should have powered by an ironless rotor, low inertia motor, which provides nearly vibration-free power.
8. Maximum Protection Class: Class 1 Degree of Protection Against Electrical Shock: Type BF II.
9. HANDPIECE
 - a. Vibration and Shock: Standard Commercial Practice
 - b. Nominal Speed: 4,500–5,500 cycles/minute