



Running Contract Details	
Equipment Name	Mayo stand
Running Contract Valid Till	27-02-2026
Tender Ref No	KMSCL/EP/T516/1214B/2023
Tendered Quantity	150
Supplier Name	M/s HOSPITAL BEDS
GST No	32AAXPF8778M1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
NALLALAM MINI INDUSTRIAL ESTATE MODERN INDUSTRIAL ESTATE NALLAM CALICUT-673027	Contact Person	NALINAKSHAN
	Phone	
	Mobile No	9495863827,4952422335
	Email	hospital.beds@hotmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Mayo stand <i>Model & Make : HOSPITAL BEDS</i>	4130 Incl.GST :18%	289.1	4419.1
		4130	289.1	4419.1

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Mayo stand

1. Should be completely stainless Steel Gr-304 matt finished.
2. Should have a stainless steel top of grade SS 304 gauge 18.
3. Dimensions should be:
 - a. Length of the Tray: 550 mm
 - b. Breadth of the Tray : 400 mm
 - c. Depth of the Tray: 50 mm
 - d. Height adjustable: 750 to 1250 mm
4. Should have 4 caster ball bearing swivel wheels of 50 cm diameter and are non-marking
5. Should be rust proof and corrosion resistant

Note:

1. If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid
2. Warranty exclusions if any shall be discussed at the time of prebid meeting else the tender condition as per clause 6.31.20 shall prevail