



Running Contract Details	
Equipment Name	Holter Monitor Model C
Running Contract Valid Till	07-08-2025
Tender Ref No	KMSCL/EP/T489(R)/160C/2023
Tendered Quantity	5
Supplier Name	M/s Schiller Healthcare India Pvt. Ltd
GST No	34AADCS5091B1ZN
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Schiller Healthcare India Pvt Ltd 37/1013 - 2nd Floor - Unidec House Fathima Church Road Elamkulam - 682 020	Contact Person	Mr.Stijo Mathew, Mr. P.S Vidhyashankar
	Phone	04842203264
	Mobile No	09846706787, 09383620520
	Email	alok@schillerindia.com, Mail@schillerindia.com, shinoj@schillerindia.com, kishorekumar@schillerindia.com, stijo@schillerindia.com, vidyashankar@schillerindia.com, mail@schillerindia.com, kishorekumar@schillerindia.com

Item-wise Price Details							
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total			
1	Holter software <i>Model & Make : Medilog AR/M/s. Schiller AG</i>	221200 Incl.GST :12%	16313.5	237513.5			
2	Recorders & accessories	207090 Incl.GST :18%	14496.3	221586.3			
3	Computer	129564 Incl.GST :18%	9069.48	138633.48			
		557854	39879.28	597733.28			
Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Holter software							
Labour	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00	10,000.00
Comprehensive	25,200.00	25,200.00	25,200.00	25,200.00	25,200.00	25,200.00	25,200.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Holter software

Equipment: Holter Monitor Model C

1. PC Supported System with following specifications
 1. Intel I5 processor
 2. Windows licensed inbuilt OS
 3. 8 GB RAM
 4. 1 TB HDD
 5. Monitor 21 inch or more
 6. Computer table
2. Saved data should be read using integrated reading device
3. Three channel ECG recording facility
4. Supporting hardware with suitable software and laser printer
5. Adult patients
6. 5 or 7 lead cable
7. Automatic cable type detection
8. Light weight
9. 24hrs, 48hrs, 7 days
10. True signal quality check with amplitude indication
11. QT/ST Analysis
12. LCD/LED Screen
13. Shock and splash-proof design
14. Suppress artifacts
15. Available scroll function for the over view (Super page) of the ECG
16. Should be able to reedit individual form classes
17. Should be able to consolidate individual form classes
18. System should be capable of analyzing various arrhythmias like ventricular ectopics, supraventricular ectopic, ventricular tachycardia, ventricular fibrillation, supraventricular tachycardia, atrial fibrillation, sinus pause
19. Operator should be able to edit and reclassify beats and arrhythmias

20. ST segment analysis should be available for all three channels
21. Any ECG print out should be possible
22. Report format should include covering page, arrhythmia analysis report, ST segment analysis, automatically and manually selected ECG strips
23. Should provide rechargeable battery and charger for recorders. Each recorder should be provided with 4 batteries and one charger

Note: If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid