



Running Contract Details	
Equipment Name	Self-Retaining Brain Retractor
Running Contract Valid Till	19-01-2023
Tender Ref No	KMSCL/EP/T356/1521/2019
Tendered Quantity	10
Supplier Name	M/s South India Surgical Co Ltd
GST No	32AAACS5091E1ZO
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
New No.117 Old No. 65 Wallajah Road Chennai-600002	Contact Person	
	Phone	
	Mobile No	
	Email	sisco@sisco.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Cranial Stabilization System with all accessories <i>Model & Make : DORO HEADREST SYSTEM,SISCO LEYLA RETRACTOR/PMI Surgical Germany</i>	880280 Incl.GST :18%	61619.6	941899.6
2	Leyla Retractor with all accessories <i>Model & Make : SISCO</i>	88883.2 Incl.GST :12%	6555.14	95438.34
		969163.2	68174.74	1037337.94

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Cranial Stabilization System with all accessories							
Labour	0.00	0.00	0.00	0.00	0.00	0.00	0.00
Comprehensive	37,500.00	37,500.00	37,500.00	37,500.00	37,500.00	37,500.00	37,500.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Cranial Stabilization System with all accessories

1. It should have support/mounting device that facilitates attachments skull clamps and head rest systems to the or table
2. It should have 33% more weight and force bearing capacity
3. It should have self locking levers for added safety
4. It should have it should have swivel adaptor which attaches the base unit to provide optimal flexibility in positioning via 360 deg. rotation
5. It should form a adaptor for head rests and skull clamps and base units
6. The skull clamp should provide visual indication of clamping force on all three skull pins , with the ability to adjust the force on all three skull pins
7. Star bursts for attaching the neuro navigation adaptor should present in the side of the skull clamp
8. Total length of the skull clamp stem should be more than 21 cm
9. Width of the base of the skull clamp should be more than 31 cm
10. Specifications of Skull Clamp Unit consisting of
 - a. QR3Skull Clamp – 1No
 - b. Adult reusable skull pin – 3 Nos
 - c. Child reusable skull pin – 3 Nos
 - d. Adjustable Base Unit – 1 No
 - e. Swivel Adaptor – 1 No
 - f. Side Rail OR Table Adaptor (Indian) – 1 No
 - g. Adult Horseshoe Headrest – 1 No
11. Specifications of Leyla Retractor consisting of
 - a. Yasargil Self Retaining Retractor, Table Mount, Double Arm - 1 No.
 - b. Ball & Socket Joint for Fixing Holding Rod -1 No.
 - c. Holding Rod for Ball & Socket -1No
 - d. Coupling Head -1 pair
 - e. Flexible Arm -2 Nos.
 - f. Support for Flat Brain Spatulas with Round Shaft -1 No.
 - g. Support for Flat Brain Spatulas -1 No.
 - h. Fixation Base for Skull Mounting Holding One Flexible Arm -1 No.