



Running Contract Details	
Equipment Name	Orthopedic Plier cum wire cutter (SS)
Running Contract Valid Till	31-07-2025
Tender Ref No	KMSCL/EP/T489/1529F/2023
Tendered Quantity	15
Supplier Name	M/s Bions Medical Systems Pvt Ltd
GST No	32AACCB4877B1Z1
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	5 Years

Supplier`s Details		
Address	Contact Details	
Address:53/333 A1 Arden Ivy Court Opp.Reliance Fresh Off S.C. Bose Road Vytila-Kochi-682019	Contact Person	JOMY THOMAS
	Phone	0484-2969622/33/44
	Mobile No	9567860500
	Email	customercare@bionsmedicalsyste.ms.com, jomy.thomas@bionsmedicalsyste.ms.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Orthopedic Plier cum wire cutter (SS)</b> <i>Model &amp; Make : Manual Instruments For Orthopedic Surgery/Johnson&amp; Johnson Pvt. Ltd.</i>	328825.35 Incl.GST :5%	25867.59	354692.94
		<b>328825.35</b>	<b>25867.59</b>	<b>354692.94</b>

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5.00% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly

to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Orthopedic Plier cum wire cutter (SS)**

#### **Item Name: Orthopedic Plier cum Wire Cutter (SS)**

1. Should have rod cutter Swiss made.
2. Bending pliers for plates 2.4-4.0 L 230
3. Parallel pliers flat nose L160

#### **Specification:**

1. AO approved design and imported bolt cutter swiss made
2. AO approved design and imported bending pliers for plates 2.4-4.0 L230
3. AO approved design and imported parallel pliers flat nose L150

#### **Note:**

1. The bidder shall offer all the items mentioned in the technical Specification, failing which the offered bid will not be consider
2. All instruments should be corrosion resistant
3. All the instruments should be rust proof
4. All the instruments should be autoclavable
5. All should have standard 5 yr warranty from manufacturers
6. Should be made of high quality medical grade stainless steel. All cutting instruments shall be SS grade 410/420.

**Note:** If CDSCO (Central Drugs Standard Control Organization) certification is required for the import and marketing of the equipment, then the same shall be submitted along with the technical bid